

CERTIFICATE OF ANALYSIS

Product Name: Q-Plex™ Human Cytokine Panel 1 (6-Plex)


Order No. 21505

REF Part No. 115233HU

LOT Lot No. HCP1K200518 17653 001

Exp Date: 2023-MAR-21

Software Product Code: HA1M200518-P16P



Reconstitute Calibrator with 650 µL of Sample Diluent

Component	REF Part No.	LOT Lot No.	Exp. Date
Plate	115252HU	HCP1P210809	2023-AUG-09
Sample Diluent	101104HU	HQ6S220107	2024-JAN-07
Calibrator	102655HU	HA1M200518	2024-APR-04
Competitor	N/A	N/A	N/A
Detection	110956HU_01	HCSDQ6220330	2024-MAR-30
Streptavidin HRP	101173GR	QNSN210921	2023-MAR-21
Substrate A	101159GR	QNSA220310	2024-MAR-10
Substrate B+	101120GR	QNSB+220310	2024-MAR-10
Wash Buffer	101158GR	QNSW220120	2024-JAN-20

Spot #	Analyte	Units	Cal Conc	ULOQ	LLOQ	LLD
1	IL-1b	pg/mL	11800	11800	16.19	8.09
2	IL-6	pg/mL	3100	3100	4.25	2.13
3	TNFa	pg/mL	5400	5400	7.41	3.70
4	IFNy	pg/mL	8000	8000	10.97	5.49
5	IL-1a	pg/mL	3900	3900	5.35	2.67
6	IL-10	pg/mL	6800	6800	9.33	4.66
7	Ref	N/A	N/A	N/A	N/A	N/A

Quality Release:

[Signature]
for Research Use Only

101896GR 2.0