

Q-Plex™ ARRAY

CERTIFICATE OF ANALYSIS

Product Name: Q-Plex™ Human Cytokine Pa
1 (6-Plex)


Order No. 21733

REF Part No. 115233HU

LOT Lot No. HCP1K200518 17877 001

EXP Exp Date: 2023-MAR-01

Software Product Code: **HA1M200518-P16P**



Reconstitute Calibrator with 650 µL of Sample Diluent

Component	REF Part No.	LOT Lot No.	EXP Exp. Date
Plate	115252HU	HCP1P211101	2023-NOV-01
Sample Diluent	101104HU	HQ6S210301	2023-MAR-01
Calibrator	102655HU	HA1M200518	2024-MAY-23
Competitor	N/A	N/A	N/A
Detection	110956HU_01	HCSDQ6220407	2024-APR-07
Streptavidin HRP	101173GR	QNSN220111	2023-JUL-11
Substrate A	101159GR	QNSA220309	2024-MAR-09
Substrate B+	101120GR	QNSB+220309	2024-MAR-09
Wash Buffer	101158GR	QNSW211201	2023-DEC-01

Spot #	Analyte	Units	Cal Conc	ULOQ	LLOQ	LLD
1	IL-1b	pg/mL	11800	3900	5.35	2.67
2	IL-6	pg/mL	3100	11800	16.19	8.09
3	TNFa	pg/mL	5400	3100	4.25	2.13
4	IFNy	pg/mL	8000	6800	9.33	4.66
5	IL-1a	pg/mL	3900	8000	10.97	5.49
6	IL-10	pg/mL	6800	5400	7.41	3.70
7	REF	N/A	N/A	N/A	N/A	N/A

Quality Release:

[Signature]
for Research Use Only

101896GR 2.0