


**CERTIFICATE OF ANALYSIS
AND PRODUCT CARD**

Product Name: Q-Plex™ Human Micronutrient v2 (7-Plex)
 Order No. 20072
 Part No. 355149HU
 Lot No. HMTK210930 17910 002
 Exp Date: 2023-JAN-05



Software Product Code: HMTMA210930-V2

**Reconstitute Calibrator with 270 µL of Sample Diluent
and 30 µL of Calibrator Additive**

Component	Part No.	Lot No.	Exp. Date
Plate	355152HU	HMT220524	2023-NOV-24
Sample Diluent	355157HU	HMTS220517	2023-NOV-17
Calibrator	355155HU	HMTM210930	2023-DEC-01
Calibrator Additive	355142HU	HMTA210930	2023-MAR-30
Competitor	355190HU	HMTC210923	2023-DEC-01
Detection	355156HU	HMTD220525	2023-NOV-25
Streptavidin HRP	101173GR	QNSN210921	2023-MAR-21
Substrate A	101159GR	QNSA210105	2023-JAN-05
Substrate B2	101128GR	QNSB2210105	2023-JAN-05
Wash Buffer	101158GR	QNSW220103	2024-JAN-03

Spot #	Analyte	Units	Cal Conc	ULOQ	LLOQ	LLD
1	Ferritin	µg/L	27.12	20.99	0.029	0.014
2	*AGP	g/L	0.49	0.49	0.00067	0.00034
3	Tg	µg/L	15.32	15.32	0.021	0.011
4	sTfR	mg/L	3.12	4.16	0.0057	0.0029
5	REF	N/A	N/A	N/A	N/A	N/A
6	CRP	mg/L	0.99	0.17	0.0001	0.00059
7	HRP2	µg/L	1.59	1.59	0.0100	0.0051
8	*RBP4	µmol/L	1.30	1.30	0.0053	0.0027

*Competitive Assay

Quality Release:

[Signature]
for Research Use Only

101896GR 2.0