


Q-Plex™ ARRAY

CERTIFICATE OF ANALYSIS

Product Name: Q-Plex™ Human Custom (4-Plex)
 Order No. 21924
REF Part No. 107749GR
LOT Lot No. HCUK200518 18171 001
 Exp Date: 2023-DEC-01
 Software
 Product Code: HA1M200518-LD4



Reconstitute Calibrator 1 with 650 µL of Sample Diluent
 Reconstitute Calibrator 2 with 400 µL of Calibrator 1

Component	Part No.	Lot No.	Exp. Date
Plate	101852GR	HCUP220726	2024-JUL-26
Sample Diluent	101104HU	HQ6S220107	2024-JAN-07
Calibrator 1	102655HU	HA1M200518	2024-AUG-02
Calibrator 2	102855HU	HA2M200303	2024-AUG-02
Detection	101856GR	HCUD220727	2024-JUL-27
Streptavidin HRP	101173GR	QNSN220711	2024-JAN-11
Substrate A	101159GR	QNSA220620	2024-JUN-20
Substrate B+	101120GR	QNSB+220620	2024-JUN-20
Wash Buffer	101158GR	QNSW211201	2023-DEC-01

Spot #	Analyte	Units	Cal Conc.	ULOQ	LLOQ	LLD
1	IL-12p40	pg/mL	2600	2600	3.57	1.78
2	IL-12p70	pg/mL	2800	2800	3.84	1.92
3	IL-23	pg/mL	42000	42000	57.61	28.81
4	IL-27	pg/mL	238200	238200	326.75	163.37
5	REF	N/A	N/A	N/A	N/A	N/A

Quality Release:

for Research Use Only

101896GR 2.0