


CERTIFICATE OF ANALYSIS

Product Name: Q-Plex™ Human Cytokine Panel 1 (4-Plex)
Order No. 22029
REF Part No. 115433HU
LOT Lot No. HCP4K200518 18224 002
Exp Date: 2023-MAY-19



Software Product Code: HA1M200518-P4Q1

Reconstitute Calibrator with 650 µL of Sample Diluent

Component	REF Part No.	LOT Lot No.	Exp. Date
Plate	115452HU	HCP4P210916	2023-SEP-16
Sample Diluent	101104HU	HQ6S210519	2023-MAY-19
Calibrator	102655HU	HA1M200518	2024-AUG-15
Competitor	N/A	N/A	N/A
Detection	110956HU_01	HCSDQ6220407	2024-APR-07
Streptavidin HRP	101173GR	QNSN220715	2024-JAN-15
Substrate A	101159GR	QNSA220713	2024-JUL-13
Substrate B+	101120GR	QNSB+220713	2024-JUL-13
Wash Buffer	101158GR	QNSW220615	2024-JUN-15

Spot #	Analyte	Units	Cal Conc	ULOQ	LLOQ	LLD
1	IL-6	pg/mL	3100	3100	4.25	2.13
2	TNFa	pg/mL	5400	5400	7.41	3.70
3	IFNy	pg/mL	8000	8000	10.97	5.49
4	IL-10	pg/mL	6800	6800	9.33	4.66
5	REF	N/A	N/A	N/A	N/A	N/A

Quality Release: *Sophia Adams*
for Research Use Only

101896GR 2.0