


CERTIFICATE OF ANALYSIS

Product Name: Q-Plex™ Human Cytokine
Panel 1 (6-Plex)


Order No. 22116

REF Part No. 115233HU


LOT Lot No. HCP1K200518 18363 001

 Exp Date: 2023-MAY-19

Software Product Code: HA1M200518-P16P



Reconstitute Calibrator with 650 µL of Sample Diluent

Component	REF Part No.	LOT Lot No.	 Exp. Date
Plate	115252HU	HCP1P211101	2023-NOV-01
Sample Diluent	101104HU	HQ6S210519	2023-MAY-19
Calibrator	102655HU	HA1M200518	2024-SEP-07
Competitor	N/A	N/A	N/A
Detection	110956HU_01	HCSDQ6220407	2024-APR-07
Streptavidin HRP	101173GR	QNSN220715	2024-JAN-15
Substrate A	101159GR	QNSA220713	2024-JUL-13
Substrate B+	101120GR	QNSB+220713	2024-JUL-13
Wash Buffer	101158GR	QNSW220803	2024-AUG-03

Spot #	Analyte	Units	Cal Conc	ULOQ	LLOQ	LLD
1	IL-1b	pg/mL	11800	11800	16.19	8.09
2	IL-6	pg/mL	3100	3100	4.25	2.13
3	TNFa	pg/mL	5400	5400	7.41	3.70
4	IFNy	pg/mL	8000	8000	10.97	5.49
5	IL-1a	pg/mL	3900	3900	5.35	2.67
6	IL-10	pg/mL	6800	6800	9.33	4.66
7	Ref	N/A	N/A	N/A	N/A	N/A

Quality Release:



101896GR 2.0

for Research Use Only