

101896GR - 4.1



Quansys Biosciences
365 N 600 W Logan, UT 84321
(435) 752-0531


CERTIFICATE OF ANALYSIS

Q-Plex™ Human Cytokine Panel 2 (6-Plex)
Software Product Code: HA1M200518-P26P

REF Part Number: 115333HU
LOT Lot Number: HCP3K200518 18547 001
Expiration Date: 2023-MAY-19

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	115352HU	HCP3P221014	N/A	2024-OCT-14
Sample Diluent	101104HU	HQ6S210519	N/A	2023-MAY-19
Calibrator Diluent	N/A	N/A	N/A	N/A
Assay Diluent	N/A	N/A	N/A	N/A
Competitor	N/A	N/A	N/A	N/A
Calibrator	102655HU	HA1M200518	650 µL	2024-OCT-17
Calibrator Additive	N/A	N/A	N/A	N/A
Controls	N/A	N/A	N/A	N/A
Detection	110956HU 01	HCSDQ6220407	N/A	2024-APR-07
Streptavidin	101173GR	QNSN220715	N/A	2024-JAN-15
Substrate A	101159GR	QNSA220713	N/A	2024-JUL-13
Substrate B+	101120GR	QNSB+220713	N/A	2024-JUL-13
Wash Buffer	101158GR	QNSW211201	N/A	2023-DEC-01

Spot #	Analyte	Units	Cal Conc	ULOQ	LLOQ	LLD
1	IL-6	pg/mL	3100	3100	4.25	2.13
2	IL-8	pg/mL	1800	1800	2.47	1.23
3	TNFα	pg/mL	5400	5400	7.41	3.70
4	IFNγ	pg/mL	8000	8000	10.97	5.49
5	IL-4	pg/mL	1600	1600	2.19	1.10
6	IL-10	pg/mL	6800	6800	9.33	4.66
7	REF	N/A	N/A	N/A	N/A	N/A



Q-Plex™ ARRAYS

QC Release: Stacey Paulsen
QA Release: J. W. [Signature]
for Research Use Only

Date: 20 OCT 22
Date: 20 OCT 22