


CERTIFICATE OF ANALYSIS

Q-Plex™ Human Pneumococcal IgG (9-Plex)
Software Product Code: HPNM007SPWHO9

REF Part Number: 481649HU
LOT Lot Number: HPNK220809 18491 001
Expiration Date: 2023-AUG-18

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	481652HU	HPNP221025	N/A	2023-OCT-25
Sample Diluent	481157HU	HPNS220818	N/A	2023-AUG-18
Calibrator Diluent	N/A	N/A	N/A	N/A
Assay Diluent	N/A	N/A	N/A	N/A
Competitor	N/A	N/A	N/A	N/A
Calibrator	481155HU	HPNM220809	500 µL	2023-NOV-03
Calibrator Additive	N/A	N/A	N/A	N/A
Controls	N/A	N/A	N/A	N/A
Detection	481156HU	HPND221027	N/A	2023-OCT-27
Streptavidin HRP	N/A	N/A	N/A	N/A
Substrate A	101159GR	QNSA220825	N/A	2024-AUG-25
Substrate B	101120GR	QNSB+220825	N/A	2024-AUG-25
Wash Buffer	101158GR	QNSW211013	N/A	2023-OCT-13

Spot #	Analvte	Units	Cal Conc	ULOQ	LLOQ	LLD
1	2	µg/mL	2.46	0.434	0.00078	0.00039
2	10A	µg/mL	1.30	0.0567	0.000416	0.000208
3	11A	µg/mL	0.51	0.091	0.000156	0.000078
4	15B	µg/mL	1.69	0.294	0.000533	0.0002665
5	17F	µg/mL	0.85	0.14	0.001079	0.0005390
6	19A	µg/mL	1.39	0.245	0.00182	0.00091
7	20	µg/mL	1.05	0.182	0.000338	0.000169
8	22F	µg/mL	0.95	0.168	0.001209	0.0006045
9	33F	µg/mL	1.07	0.0469	0.000338	0.000169
10	CWPS	N/A	N/A	N/A	N/A	N/A
11	DK anti-Hu	N/A	N/A	N/A	N/A	N/A



Q-Plex™ ARRAYS

QC Release: Sylvia Adams
QA Release: J. Wa
for Research Use Only

Date: 03NOV22
Date: 03NOV22