



CERTIFICATE OF ANALYSIS

Q-Plex™ Human MMP (6-Plex)
 Software Product Code: MMPM220518

REF Part Number: 340949HU
LOT Lot Number: MMPK220518 18631 001
 Expiration Date: 2023-JUL-16

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	340952HU	MMPP220630	N/A	2024-JUN-30
Sample Diluent	340957HU	MMPS220124	N/A	2024-JAN-24
Calibrator Diluent	N/A	N/A	N/A	N/A
Assay Diluent	340983HU	MMPV220124	N/A	2024-JAN-24
Competitor	N/A	N/A	N/A	N/A
Calibrator	340955HU	MMPM220518	300 µL	2024-NOV-09
Calibrator Additive	N/A	N/A	N/A	N/A
Controls	N/A	N/A	N/A	N/A
Detection	340956HU	MMPD210716	N/A	2023-JUL-16
Streptavidin HRP	101173GR	QNSN220715	N/A	2024-JAN-15
Substrate A	101159GR	QNSA220825	N/A	2024-AUG-25
Substrate B+	101120GR	QNSB+220825	N/A	2024-AUG-25
Wash Buffer	101158GR	QNSW220909	N/A	2024-SEP-09

Spot #	Analyte	Units	Cal Conc	ULOQ	LLOQ	LLD
1	MMP 1	pg/mL	12563	12563	51.70	25.85
2	MMP 2	pg/mL	85197	85197	350.60	175.30
3	MMP 3	pg/mL	4579	4579	18.84	9.42
4	MMP 7	pg/mL	1745	1745	7.18	3.59
5	MMP 9	pg/mL	44914	44914	184.83	92.42
6	MMP 13	pg/mL	32943	32943	135.57	67.78
7	REF	N/A	N/A	N/A	N/A	N/A



Q-Plex™ ARRAYS

QC Release:

Stacey Prusbery

Date:

09 NOV 22

QA Release:

Sophia Adams

Date:

09 NOV 22

for Research Use Only