

101896GR - 4.2



Quansys Biosciences
365 N 600 W Logan, UT 84321
(435) 752-0531


CERTIFICATE OF ANALYSIS

Q-Plex™ Human Custom 7-Plex
Software Product Code: HA1M200316-UC

REF Part Number: 107749GR
LOT Lot Number: HCUK200316 18720 001
Expiration Date: 2023-MAY-19

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	101852GR	HCUP211014	N/A	2023-OCT-14
Sample Diluent	101104HU	HQ6S210519	N/A	2023-MAY-19
Calibrator Diluent	N/A	N/A	N/A	N/A
Assay Diluent	N/A	N/A	N/A	N/A
Calibrator 1	102655HU	HA1M200316	600 µL	2024-DEC-01
Calibrator 2	102855HU	HA2M200303	400 µL	2024-DEC-01
Calibrator Additive	N/A	N/A	N/A	N/A
Controls	N/A	N/A	N/A	N/A
Detection	101856GR	HCUD211019	N/A	2023-OCT-19
Streptavidin	101173GR	QNSN220715	N/A	2024-JAN-15
Substrate A	101159GR	QNSA220825	N/A	2024-AUG-25
Substrate B+	101120GR	QNSB+220825	N/A	2024-AUG-25
Wash Buffer	101158GR	QNSW211013	N/A	2023-OCT-13

Spot #	Analyte	Units	Cal Conc	ULOQ	LLOQ	LLD
1	Eotaxin	pg/mL	1700	1700	2.33	1.17
2	IL-1Ra	pg/mL	7930	7930	10.88	5.44
3	IL-6	pg/mL	2900	2900	3.98	1.99
4	IL-8	pg/mL	1900	1900	2.61	1.30
5	IL-10	pg/mL	7600	7600	10.43	5.21
6	IP-10	pg/mL	3000	1000	4.12	2.06
7	MCP-1	pg/mL	2000	667	2.74	1.37
8	REF	N/A	N/A	N/A	N/A	N/A



Q-Plex™ ARRAYS

QC Release: Sophia Adams Date: 01 DEC 22
QA Release: [Signature] Date: 01 DEC 22

for Research Use Only