


101896GR - 5.0

Quansys Biosciences
365 N 600 W Logan, UT 84321
(435) 752-0531**CERTIFICATE OF ANALYSIS**Q-Plex™ Human Innate Immunity (7-Plex)
Software Product Code: HA1M200518-IN

REF Part Number: 111233HU
LOT Lot Number: HINK200518 18705 001
Expiration Date: 2024-JAN-07

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	111252HU	HINP221201	N/A	2024-DEC-01
Sample Diluent	101104HU	HQ6S220107	N/A	2024-JAN-07
Calibrator Diluent	N/A	N/A	N/A	N/A
Assay Diluent	N/A	N/A	N/A	N/A
Calibrator 1	102655HU	HA1M200518	650 µL	2024-DEC-19
Calibrator 2	102855HU	HA2M200303	400 µL	2024-DEC-19
Calibrator Additive	N/A	N/A	N/A	N/A
Controls	N/A	N/A	N/A	N/A
Detection	101856GR	HCUD221209	6 mL	2024-DEC-09
Streptavidin HRP	101173GR	QNSN221024	N/A	2024-APR-24
Substrate A	101159GR	QNSA221020	N/A	2024-OCT-20
Substrate B+	101120GR	QNSB+221020	N/A	2024-OCT-20
Wash Buffer	101158GR	QNSW220909	N/A	2024-SEP-09

Spot #	Analyte	Units	Cal Conc	ULOQ	LLOQ	LLD
1	IL-1α	pg/mL	3900	3900	5.35	2.67
2	IL-1β	pg/mL	11800	11800	16.19	8.09
3	IFNα	pg/mL	380	380	0.52	0.26
4	IFNβ	pg/mL	21700	21700	29.77	14.88
5	IFNγ	pg/mL	8000	8000	10.97	5.49
6	GMCSF	pg/mL	6000	6000	8.23	4.12
7	IL-18	pg/mL	235	235	0.32	0.16
8	REF	N/A	N/A	N/A	N/A	N/A



Q-Plex™ ARRAYS

QC Release:

Sophie Adams

Date:

19 DEC 22

QA Release:

J. Wa

Date:

19 DEC 22

for Research Use Only