

CERTIFICATE OF ANALYSIS


Q-Plex™ Mouse Cytokine Panel 3 HS (4-Plex)
Software Product Code: MHSM220208-P34PR



REF Part Number: 111249MS
LOT Lot Number: MCP3K220208 18716 001
Expiration Date: 2023-SEP-26

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	111252MS	MCP3P221201	N/A	2023-DEC-01
Sample Diluent	112457MS	MHSS221021	N/A	2023-OCT-21
Calibrator Diluent	N/A	N/A	N/A	N/A
Assay Diluent	N/A	N/A	N/A	N/A
Competitor	N/A	N/A	N/A	N/A
Calibrator	112455MS	MHSM220208	650 µL	2023-DEC-06
Calibrator Additive	N/A	N/A	N/A	N/A
Controls	N/A	N/A	N/A	N/A
Detection	112456MS	MHSD220926	6 mL	2023-SEP-26
Streptavidin HRP	101173GR	QNSN221024	N/A	2024-APR-24
Substrate A	101159GR	QNSA221020	N/A	2024-OCT-20
Substrate B+	101120GR	QNSB+221020	N/A	2024-OCT-20
Wash Buffer	101158GR	QNSW221123	N/A	2024-NOV-23

Spot #	Analyte	Cal Conc	Units	ULOQ	LLOQ	LLD
1	IL-2	170	pg/mL	270	0.37	0.19
2	IL-4	150	pg/mL	150	0.21	0.10
3	IL-6	1180	pg/mL	1180	1.62	0.81
4	TNFα	400	pg/mL	400	0.75	0.38
5	REF	N/A	N/A	N/A	N/A	N/A



Q-Plex™ ARRAYS

QC Release: [Signature]

Date: 21SEP23

QA Release: [Signature]

Date: 21SEP23

for Research Use Only