


CERTIFICATE OF ANALYSIS

Q-Plex™ NHP Custom 10-Plex
Software Product Code: NHPM220606-LC10

REF Part Number: 107749GR
Lot Lot Number: NHPCUK220606 18702 001
Expiration Date: 2023-NOV-03

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	101852GR	NHPCUP221110	N/A	2023-NOV-10
Sample Diluent	101104HU	HQ6S220107	N/A	2024-JAN-07
Calibrator Diluent	N/A	N/A	N/A	N/A
Assay Diluent	N/A	N/A	N/A	N/A
Calibrator 1	109255HU	NHPM220606	500 µL	2023-NOV-29
Calibrator 2	101855GR	NHPCUM221027	400 µL	2023-NOV-29
Calibrator Additive	N/A	N/A	N/A	N/A
Controls	N/A	N/A	N/A	N/A
Detection	101856GR	NHPCUD221103	6 mL	2023-NOV-03
Streptavidin HRP	101173GR	QNSN221024	N/A	2024-APR-24
Substrate A	101159GR	QNSA221020	N/A	2024-OCT-20
Substrate B+	101120GR	QNSB+221020	N/A	2024-OCT-20
Wash Buffer	101158GR	QNSW220930	N/A	2024-SEP-30

Spot #	Analyte	Units	Cal Conc	ULOQ	LLOQ	LLD
1	IL-1Ra	pg/mL	6490	6490	8.90	4.45
2	MIP-1α	pg/mL	1800	600	2.47	1.23
3	MIP-1β	pg/mL	5020	5020	6.89	3.44
4	IL-15	pg/mL	3100	3100	4.25	2.13
5	VEGF	pg/mL	2050	2050	2.81	1.41
6	IL-12p40	pg/mL	2630	2630	3.61	1.80
7	GCSF	pg/mL	5110	5110	7.01	3.50
8	IL-18	pg/mL	186	186	0.26	0.13
9	IFNγ	pg/mL	820	820	1.12	0.56
10	TNFα	pg/mL	190	190	0.26	0.13
11	REF	N/A	N/A	N/A	N/A	N/A



Q-Plex™ ARRAYS

QC Release: *Haley Prober*
QA Release: *J. Wa*
for Research Use Only

Date: 01 DEC 22
Date: 01 DEC 22