


**CERTIFICATE OF ANALYSIS**

Q-Plex™ Human Cytokine Panel 2 (6-Plex)  
Software Product Code: HA1M200518-P26P

**REF** Part Number: 115333HU  
**LOT** Lot Number: HCP3K200518 18970 001  
Expiration Date: 2024-APR-07

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	115352HU	HCP3P221014	N/A	2024-OCT-14
Sample Diluent	101104HU	HQ6S220909	N/A	2024-SEP-09
Calibrator Diluent	N/A	N/A	N/A	N/A
Assay Diluent	N/A	N/A	N/A	N/A
Competitor	N/A	N/A	N/A	N/A
Calibrator	102655HU	HA1M200518	650 µL	2025-FEB-06
Calibrator Additive	N/A	N/A	N/A	N/A
Controls	N/A	N/A	N/A	N/A
Detection	110956HU_01	HCSDQ6220407	N/A	2024-APR-07
Streptavidin HRP	101173GR	QNSN221228	N/A	2024-JUN-28
Substrate A	101159GR	QNSA221020	N/A	2024-OCT-20
Substrate B+	101120GR	QNSB+221020	N/A	2024-OCT-20
Wash Buffer	101158GR	QNSW221123	N/A	2024-NOV-23

Spot #	Analyte	Units	Cal Conc	ULOQ	LLOQ	LLD
1	IL-6	pg/mL	3100	3100	4.25	2.13
2	IL-8	pg/mL	1800	1800	2.47	1.23
3	TNFα	pg/mL	5400	5400	7.41	3.70
4	IFNγ	pg/mL	8000	8000	10.97	5.49
5	IL-4	pg/mL	1600	1600	2.19	1.10
6	IL-10	pg/mL	6800	6800	9.33	4.66
7	REF	N/A	N/A	N/A	N/A	N/A



**Q-Plex™** ARRAYS

QC Release: Sophia Adams

Date: 06 FEB 23

QA Release: [Signature]

Date: 06 FEB 23

for Research Use Only