

101896GR - 5.0



Quansys Biosciences
365 N 600 W Logan, UT 84321
(435) 752-0531


CERTIFICATE OF ANALYSIS

Q-Plex™ Mouse Custom 3-Plex
Software Product Code: MA1M191126-CC3

REF Part Number: 107749GR
LOT Lot Number: MCUK191126 19070 001
Expiration Date: 2023-NOV-20

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	101852GR	MCUP220609	N/A	2024-JUN-09
Sample Diluent	110957MS	MCSS220607	N/A	2024-JUN-07
Calibrator Diluent	N/A	N/A	N/A	N/A
Assay Diluent	N/A	N/A	N/A	N/A
Competitor	N/A	N/A	N/A	N/A
Calibrator	102655MS	MA1M191126	550 µL	2024-SEP-12
Calibrator Additive	N/A	N/A	N/A	N/A
Controls	N/A	N/A	N/A	N/A
Detection	101856GR	MCUD220610	N/A	2024-JUN-10
Streptavidin HRP	101173GR	QNSN220520	N/A	2023-NOV-20
Substrate A	101159GR	QNSA211217	N/A	2023-DEC-17
Substrate B	101120BR	QNSB+211217	N/A	2023-DEC-17
Wash Buffer	101158GR	QNSW220422	N/A	2024-APR-22

Spot #	Analyte	Units	Cal Conc	ULOQ	LLOQ	LLD
1	IL-1β	pg/mL	12950	N/A	N/A	12.41
2	IL-6	pg/mL	4800	N/A	N/A	2.90
3	TNFα	pg/mL	2980	N/A	N/A	3.40
4	REF	N/A	N/A	N/A	N/A	N/A



Q-Plex™ ARRAYS

QC Release: *[Signature]*

Date: 03 MAR 23

QA Release: *[Signature]*

Date: 03 MAR 23

for Research Use Only