

101896GR - 6.0



Quansys Biosciences
365 N 600 W Logan, UT 84321
(435) 752-0531


CERTIFICATE OF ANALYSIS

Q-Plex™ Human Pneumococcal IgG (9-Plex)
Software Product Code: HPNM007SPWHO9

REF Part Number: 481649HU
LOT Lot Number: HPNK220926 19459 001
EXP Expiration Date: 2023-AUG-16

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	481652HU	HPNP220816	N/A	2023-AUG-16
Sample	481157HU	HPNS221114	N/A	2023-NOV-14
Calibrator	N/A	N/A	N/A	N/A
Assay Diluent	N/A	N/A	N/A	N/A
Competitor	N/A	N/A	N/A	N/A
Calibrator	481155HU	HPNM220926	500 µL	2024-JUN-13
Calibrator Additive	N/A	N/A	N/A	N/A
Controls	N/A	N/A	N/A	N/A
Detection	481156HU	HPND221027	N/A	2023-OCT-27
Streptavidin	N/A	N/A	N/A	N/A
Substrate A	101159GR	QNSA230302	N/A	2025-MAR-02
Substrate B+	101120GR	QNSB+230302	N/A	2025-MAR-02
Wash Buffer	101158GR	QNSW230314	N/A	2025-MAR-14

Spot #	Analyte	Units	Cal Conc	ULOQ	LLOQ	LLD
1	2	µg/mL	2.463	0.434	0.00078	0.00039
2	10A	µg/mL	1.298	0.0567	0.00042	0.00021
3	11A	µg/mL	0.508	0.091	0.00016	0.000078
4	15B	µg/mL	1.694	0.294	0.00053	0.00027
5	17F	µg/mL	0.851	0.14	0.0011	0.00054
6	19A	µg/mL	1.387	0.245	0.0018	0.00091
7	20	µg/mL	1.047	0.182	0.00034	0.00017
8	22F	µg/mL	0.950	0.168	0.0012	0.00061
9	33F	µg/mL	1.066	0.0469	0.00034	0.00017
10	CWPS	N/A	N/A	N/A	N/A	N/A
11	REF	N/A	N/A	N/A	N/A	N/A



Q-Plex™ ARRAYS

QC Release: [Signature]
QA Release: [Signature]
for Research Use Only

Date: 14 JUN 23
Date: 14 JUN 23