

101896GR - 7.0

QUANSYS
BIOSCIENCES
CERTIFICATE OF ANALYSIS

Quansys Biosciences
365 N 600 W Logan, UT 84321
(435) 752-0531


Q-Plex™ Human Custom 23-Plex
Software Product Code: HPNM230726-WHO23



REF Part Number: 107749GR
LOT Lot Number: HCUK230726 19812 001
Expiration Date: 2024-FEB-28

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	101852GR	HCUP230822	N/A	2024-AUG-22
Sample Diluent	481157HU	HPNS230228	N/A	2024-FEB-28
Calibrator Diluent	N/A	N/A	N/A	N/A
Assay Diluent	N/A	N/A	N/A	N/A
Competitor	N/A	N/A	N/A	N/A
Calibrator	481155HU	HPNM230726	500 µL	2024-SEP-07
Calibrator Additive	N/A	N/A	N/A	N/A
Controls	N/A	N/A	N/A	N/A
Detection	481156HU	HPND230730	N/A	2024-JUL-30
Streptavidin HRP	N/A	N/A	N/A	N/A
Substrate A	101159GR	QNSA230602	N/A	2025-JUN-02
Substrate B+	101120GR	QNSB+230602	N/A	2025-JUN-02
Wash Buffer	101158GR	QNSW230615	N/A	2025-JUN-15

Spot #	Serotype	Cal Conc	Units	ULOQ	LLOQ	LLD
1	REF	N/A	N/A	N/A	N/A	N/A
2	18C	0.73	µg/mL	0.13	0.00093	0.00047
3	7F	0.83	µg/mL	0.15	0.0011	0.00053
4	14	3.80	µg/mL	0.16	0.0012	0.00060
5	9V	0.64	µg/mL	0.11	0.00079	0.00039
6	33F	1.07	µg/mL	0.047	0.00034	0.00017
7	10A	1.30	µg/mL	0.057	0.00042	0.00021
8	20	1.05	µg/mL	0.18	0.00034	0.00017
9	15B	1.69	µg/mL	0.29	0.00053	0.00027
10	4	0.33	µg/mL	0.058	0.00010	0.000052
11	2	2.46	µg/mL	0.43	0.00078	0.00039
12	11A	0.51	µg/mL	0.091	0.00016	0.000078
13	5	0.75	µg/mL	0.13	0.00024	0.00012
14	8	1.42	µg/mL	0.25	0.0018	0.00090
15	9N	0.70	µg/mL	0.12	0.00089	0.00045
16	1	0.85	µg/mL	0.15	0.00027	0.00013
17	6B	0.91	µg/mL	0.16	0.00029	0.00015
18	17F	0.85	µg/mL	0.14	0.0011	0.00054
19	23F	0.60	µg/mL	0.10	0.00076	0.00038
20	22F	0.95	µg/mL	0.17	0.0012	0.00061
21	12F	0.22	µg/mL	0.038	0.00028	0.00014
22	19F	1.46	µg/mL	0.26	0.0019	0.00094
23	19A	1.39	µg/mL	0.25	0.0018	0.00091
24	3	0.15	µg/mL	0.10	0.00019	0.000093
25	CWPS	N/A	N/A	N/A	N/A	N/A



Q-Plex™ ARRAYS

QC Release: [Signature]

Date: 07 SEP 23

QA Release: [Signature]

Date: 07 SEP 23

for Research Use Only