

101896GR - 7.0

QUANSYS 
 BIOSCIENCES
CERTIFICATE OF ANALYSIS

Quansys Biosciences
 365 N 600 W Logan, UT 84321
 (435) 752-0531

Q-Plex™ Human Angiogenesis (9-Plex)

Software Product Code: HA1M220712-AG




REF Part Number: 150233HU

LOT Lot Number: HAGK220712 20149 001


EXP Expiration Date: 2024-DEC-28

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	150252HU	HAGP231222	N/A	2025-DEC-22
Sample Diluent	101104HU	HQ6S230525	N/A	2025-MAY-25
Calibrator Diluent	N/A	N/A	N/A	N/A
Assay Diluent	N/A	N/A	N/A	N/A
Competitor	N/A	N/A	N/A	N/A
Calibrator	102655HU	HA1M220712	600 µL	2026-JAN-11
Calibrator Additive	N/A	N/A	N/A	N/A
Controls	N/A	N/A	N/A	N/A
Detection	150256HU 01	HAGDQ6240105	N/A	2026-JAN-05
Streptavidin HRP	101173GR	QNSN230628	N/A	2024-DEC-28
Substrate A	101159GR	QNSA231006	N/A	2025-MAR-13
Substrate B+	101120GR	QNSB+231006	N/A	2025-MAR-13
Wash Buffer	101158GR	QNSW230718	N/A	2025-JUL-18

Spot #	Analyte	Cal Conc	Units	ULOQ	LLOQ	LLD
1	ANG-2	8460	pg/mL	8460	11.60	5.80
2	FGF	4950	pg/mL	4950	6.79	3.40
3	HGF	15160	pg/mL	15160	20.80	10.40
4	IL-8	2080	pg/mL	2080	2.85	1.43
5	PDGF	4100	pg/mL	4100	5.62	2.81
6	TIMP-1	12840	pg/mL	12840	52.84	26.42
7	TIMP-2	19135	pg/mL	19135	26.25	13.12
8	TNFα	6020	pg/mL	6020	8.26	4.13
9	VEGF	1890	pg/mL	1890	2.59	1.30
10	REF	N/A	N/A	N/A	N/A	N/A



Q-Plex™ ARRAYS

QC Release: 

Date: 11 JAN 24

QA Release: 

Date: 11 JAN 24

for Research Use Only