

101896GR - 7.0

QUANSYS BIOSCIENCES CERTIFICATE OF ANALYSIS

Quansys Biosciences
365 N 600 W Logan, UT 84321
(435) 752-0531

Q-Plex™ Bovine Cytokine Panel 1 (5-Plex)
Software Product Code: BC1M230308




REF Part Number: 112349BV

LOT Lot Number: BC1K230308 20637 001

EXP Expiration Date: 2024-OCT-13

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	112352BV	BC1P230228	N/A	2025-FEB-28
Sample Diluent	101104HU	HQ6S230525	N/A	2025-MAY-25
Calibrator Diluent	N/A	N/A	N/A	N/A
Assay Diluent	N/A	N/A	N/A	N/A
Competitor	N/A	N/A	N/A	N/A
Calibrator	112355BV	BC1M230308	500 µL	2026-APR-09
Calibrator Additive	N/A	N/A	N/A	N/A
Controls	N/A	N/A	N/A	N/A
Detection	112356BV	BC1D230307	N/A	2025-MAR-07
Streptavidin HRP	101173GR	QNSN240103	N/A	2025-JUL-03
Substrate A	101159GR	QNSA231205	N/A	2024-OCT-13
Substrate B+	101120GR	QNSB+231205	N/A	2024-OCT-13
Wash Buffer	101158GR	QNSW240208	N/A	2026-FEB-08

Spot #	Analyte	Cal Conc	Units	ULOQ	LLOQ	LLD
1	IFN γ	1604	pg/mL	1604	2.20	1.10
2	IL-4	477	pg/mL	477	0.65	0.33
3	IL-10	5861	pg/mL	1954	8.04	4.02
4	IL-12p40	210911	pg/mL	210911	289.32	144.66
5	TNF α	3121	pg/mL	3121	12.84	6.42
6	REF	N/A	N/A	N/A	N/A	N/A



Q-Plex™ ARRAYS

QC Release: *[Signature]*

Date: 09 APR 24

QA Release: *[Signature]*

Date: 09 APR 24

for Research Use Only