

101896GR - 7.0



Quansys Biosciences
365 N 600 W Logan, UT 84321
(435) 752-0531


CERTIFICATE OF ANALYSIS

Q-Plex™ Human Pneumococcal IgG (14-Plex) Aliquoted Bulk 30 mL Diluent
Software Product Code: HPNM231127-A14P

REF Part Number: 481349HU-25
LOT Lot Number: HPNK231127 20634 001
Expiration Date: 2024-NOV-27

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	481852HU	HPNP231127	N/A	2024-NOV-27
Sample Diluent	481157HU-30	HPNS240401	N/A	2025-APR-01
Calibrator Diluent	N/A	N/A	N/A	N/A
Assay Diluent	N/A	N/A	N/A	N/A
Competitor	N/A	N/A	N/A	N/A
Calibrator	481155HU	HPNM231127	500 µL	2025-APR-08
Calibrator Additive	N/A	N/A	N/A	N/A
Controls	N/A	N/A	N/A	N/A
Detection	481156HU	HPND231128	N/A	2024-NOV-28
Streptavidin HRP	N/A	N/A	N/A	N/A
Substrate A	101159GR	QNSA231027	N/A	2025-AUG-23
Substrate B+	101120GR	QNSB+231027	N/A	2025-AUG-23
Wash Buffer	101158GR	QNSW231218	N/A	2025-DEC-18

Spot #	Analyte	Cal Conc	Units	ULOQ	LLOQ	LLD
1	1	1.20	µg/mL	0.26	0.0019	0.00095
2	3	0.36	µg/mL	0.29	0.0021	0.0011
3	4	0.72	µg/mL	0.15	0.0011	0.00060
4	5	1.81	µg/mL	0.37	0.0027	0.0014
5	6B	0.63	µg/mL	0.13	0.0010	0.00050
6	7F	2.01	µg/mL	0.39	0.011	0.0057
7	8	0.61	µg/mL	0.13	0.0038	0.0019
8	9N	0.73	µg/mL	0.15	0.0044	0.0022
9	9V	1.03	µg/mL	0.20	0.0057	0.0029
10	12F	0.36	µg/mL	0.070	0.0021	0.0011
11	14	2.26	µg/mL	0.12	0.0035	0.0018
12	18C	1.06	µg/mL	0.22	0.0062	0.0031
13	19F	4.53	µg/mL	0.92	0.027	0.014
14	23F	0.41	µg/mL	0.084	0.0025	0.0013
15	CWPS	N/A	N/A	N/A	N/A	N/A
16	REF	N/A	N/A	N/A	N/A	N/A



Q-Plex™ ARRAYS

QC Release: [Signature] Date: 08 APR 24
 QA Release: [Signature] Date: 08 APR 24

for Research Use Only