

101896GR - 7.0

**QUANSYS**  
BIOSCIENCES  
**CERTIFICATE OF ANALYSIS**

Quansys Biosciences  
365 N 600 W Logan, UT 84321  
(435) 752-0531

Q-Plex™ Human Pneumococcal IgG (14-Plex)  
Software Product Code: HPNM231211-14P

REF Part Number: 481849HUD  
LOT Lot Number: HPNK231211 20959 001  
Expiration Date: 2024-NOV-27

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	481852HU	HPNP231127	N/A	2024-NOV-27
Sample Diluent	481157HUD	HPNS240624	N/A	2025-JUN-24
Calibrator Diluent	N/A	N/A	N/A	N/A
Assay Diluent	N/A	N/A	N/A	N/A
Competitor	N/A	N/A	N/A	N/A
Calibrator	481155HU	HPNM231211	500 µL	2025-JUN-25
Calibrator Additive	N/A	N/A	N/A	N/A
Controls	N/A	N/A	N/A	N/A
Detection	481156HU	HPND231212	N/A	2024-DEC-12
Streptavidin HRP	N/A	N/A	N/A	N/A
Substrate A	101159GR	QNSA240405	N/A	2026-MAR-19
Substrate B+	101120GR	QNSB+240405	N/A	2026-MAR-19
Wash Buffer	101158GR	QNSW240514	N/A	2026-MAY-14

Spot #	Analyte	Cal Conc	Units	ULOQ	LLOQ	LLD
1	1	0.76	µg/mL	0.19	0.00019	0.000093
2	3	0.12	µg/mL	0.12	0.00012	0.000061
3	4	0.29	µg/mL	0.073	0.00007	0.000036
4	5	0.74	µg/mL	0.18	0.00018	0.000090
5	6B	0.85	µg/mL	0.21	0.00021	0.00010
6	7F	0.77	µg/mL	0.19	0.00075	0.00038
7	8	1.30	µg/mL	0.33	0.0013	0.00064
8	9N	0.68	µg/mL	0.17	0.00066	0.00033
9	9V	0.58	µg/mL	0.14	0.00056	0.00028
10	12F	0.20	µg/mL	0.050	0.00019	0.000097
11	14	3.43	µg/mL	0.21	0.00084	0.00042
12	18C	0.66	µg/mL	0.17	0.00065	0.00032
13	19F	1.24	µg/mL	0.31	0.0012	0.00060
14	23F	0.53	µg/mL	0.13	0.00052	0.00026
15	CWPS	N/A	N/A	N/A	N/A	N/A
16	REF	N/A	N/A	N/A	N/A	N/A



QC Release: [Signature]  
QA Release: [Signature]  
for Research Use Only

Date: 27 JUN 24  
Date: 27 JUN 24