


101896GR - 8.0

QUANSYS
BIOSCIENCES
CERTIFICATE OF ANALYSISQuansys Biosciences
365 N 600 W
Logan, UT 84321
(435) 752 - 0531**Q-Plex™ Human Custom 10-Plex**

REF	Part Number: 107749GR-10	Software Product Code: HA1M220712-L10 
LOT	Lot Number: HCUK220712 21430 001	
	Expiration Date: 2025-SEP-06	

Reconstitute Calibrator 1 using 600 µL of Sample Diluent
Reconstitute Calibrator 2 using 400 µL of Calibrator 1

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	101852GR	HCUP241008	N/A	2026-OCT-08
Sample Diluent	101104HU	HQ6S240125	N/A	2026-JAN-25
Calibrator 1	102655HU	HA1M220712	600 µL	2026-OCT-11
Calibrator 2	101855HU	HCUM210831	400 µL	2026-OCT-11
Detection	101856GR	HCUD240628	6 mL	2026-JUN-28
Streptavidin HRP	101173GR	QNSN240508	N/A	2025-SEP-06
Substrate A	101159GR	QNSA240906	N/A	2026-MAR-19
Substrate B+	101120GR	QNSB+240906	N/A	2026-MAR-19
Wash Buffer	101158GR	QNSW231122	N/A	2025-NOV-22

**Q-Plex™** ARRAYS

Spot #	Analyte	Cal Conc	Units	LLD	LLOQ	ULOQ
1	IFNα	385	pg/mL	0.26	0.53	385
2	IFNγ	9775	pg/mL	6.70	13.41	9775
3	IFNλ	12080	pg/mL	8.29	16.57	12080
4	IL-1β	11800	pg/mL	8.09	16.19	11800
5	IL-2	5150	pg/mL	3.53	7.06	5150
6	IL-6	2850	pg/mL	1.95	3.91	2850
7	IL-8	2080	pg/mL	1.43	2.85	2080
8	IL-10	8400	pg/mL	5.76	11.52	8400
9	IP-10	3240	pg/mL	2.22	4.44	1080
10	TNFα	6020	pg/mL	4.13	8.26	6020
11	REF	N/A	N/A	N/A	N/A	N/A

QC Release:

QA Release:

Date:

Date:

for Research Use Only