



101896GR - 8.0

QUANSYS BIOSCIENCES
CERTIFICATE OF ANALYSIS


Quansys Biosciences
 365 N 600 W
 Logan, UT 84321
 (435) 752-0531

Q-Plex™ Human Pneumococcal IgG (9-Plex)

REF	Part Number: 481649HU	Software Product Code: HPNM231211-9P 
LOT	Lot Number: HPNK231211 21781 001	
	Expiration Date: 2025-JUN-18	

Reconstitute Calibrator using 500 µL of Sample Diluent

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	481652HU	HPNP240618	N/A	2025-JUN-18
Sample Diluent	481157HU	HPNS240816	N/A	2025-AUG-16
Calibrator	481155HU	HPNM231211	500 µL	2026-JAN-17
Detection	481156HU	HPND240621	N/A	2025-JUN-21
Substrate A	101159GR	QNSA240801	N/A	2026-MAR-19
Substrate B+	101120GR	QNSB+240801	N/A	2026-MAR-19
Wash Buffer	101158GR	QNSW241030	N/A	2026-OCT-30



Q-Plex™ ARRAYS

Spot #	Analyte	Cal Conc	Units	LLD	LLOQ	ULOQ
1	2	2.25	µg/mL	0.0011	0.0022	0.56
2	10A	1.19	µg/mL	0.00058	0.0012	0.074
3	11A	0.47	µg/mL	0.00023	0.00046	0.12
4	15B	1.55	µg/mL	0.00076	0.0015	0.39
5	17F	0.73	µg/mL	0.0014	0.0028	0.18
6	19A	1.28	µg/mL	0.0025	0.0050	0.32
7	20	0.92	µg/mL	0.00045	0.00090	0.23
8	22F	0.87	µg/mL	0.0017	0.0034	0.22
9	33F	0.94	µg/mL	0.00046	0.00092	0.059
10	CWPS	N/A	N/A	N/A	N/A	N/A
11	REF	N/A	N/A	N/A	N/A	N/A

QC Release: *[Signature]* Date: 17 JAN 25
 QA Release: *[Signature]* Date: 17 JAN 25
 for Research Use Only