

101896GR - 8.1

QUANSYS
BIOSCIENCES
CERTIFICATE OF ANALYSIS


Quansys Biosciences
365 N 600 W
Logan, UT 84321
(435) 752 - 0531

Q-Plex™ Human Custom 3-Plex

REF	Part Number: 107749GR-03	Software Product Code:
LOT	Lot Number: HCUK220719 22099 001	HA1M220719-L3
	Expiration Date: 2026-MAR-19	

Reconstitute Calibrator using 650 µL of Sample Diluent

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	101852GR	HCUP250224	N/A	2027-FEB-24
Sample Diluent	101104HU	HQ6S240805	N/A	2026-AUG-05
Calibrator	102655HU	HA1M220719	650 µL	2027-MAR-27
Detection	101856GR	HCUD250321	6 mL	2027-MAR-21
Streptavidin HRP	101173GR	QNSN241105	N/A	2026-MAY-05
Substrate A	101159GR	QNSA241108	N/A	2026-MAR-19
Substrate B+	101120GR	QNSB+241108	N/A	2026-MAR-19
Wash Buffer	101158GR	QNSW250106	N/A	2027-JAN-06



Spot #	Analyte	Cal Conc	Units	LLD	LLOQ	ULOQ
1	IL-1β	11000	pg/mL	7.54	15.09	11000
2	IL-6	2710	pg/mL	1.86	3.72	2710
3	TNFα	5490	pg/mL	3.77	7.53	5490
4	REF	N/A	N/A	N/A	N/A	N/A

Q-Plex™ ARRAYS


QC Release:



Date:

27 MAR 25

QA Release:


for Research Use Only

Date:

27 MAR 25