

101896GR - 8.1

QUANSYS
 BIOSCIENCES
CERTIFICATE OF ANALYSIS

Quansys Biosciences
 365 N 600 W
 Logan, UT 84321
 (435) 752-0531

Q-Plex™ Human Pneumococcal IgG (9-Plex)

REF	Part Number: 482149HU	Software Product Code: HPNM230726-9.3 
LOT	Lot Number: HPNK230726 22123 001	
	Expiration Date: 2026-JAN-21	

Reconstitute Calibrator using 500 µL of Sample Diluent

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	482152HU	HPNP250408	N/A	2026-APR-08
Sample Diluent	481157HU	HPNS250407	N/A	2026-APR-07
Calibrator	481155HU	HPNM230726	500 µL	2026-APR-11
Detection	482156HU	HPNP250121	N/A	2026-JAN-21
Substrate A	101159GR	QNSA250227	N/A	2026-DEC-09
Substrate B+	101120GR	QNSB+250227	N/A	2026-DEC-09
Wash Buffer	101158GR	QNSW250106	N/A	2027-JAN-06



Spot #	Analyte	Cal Conc	Units	LLD	LLOQ	ULOQ
1	CWPS	N/A	N/A	N/A	N/A	N/A
2	2	2.46	µg/mL	0.0012	0.0024	0.62
3	10A	1.30	µg/mL	0.00063	0.0013	0.081
4	11A	0.51	µg/mL	0.00025	0.00050	0.13
5	15B	1.69	µg/mL	0.00083	0.0017	0.42
6	17F	0.85	µg/mL	0.0017	0.0033	0.21
7	19A	1.39	µg/mL	0.0027	0.0054	0.35
8	20	1.05	µg/mL	0.00051	0.0010	0.26
9	22F	0.95	µg/mL	0.0019	0.0037	0.24
10	33F	1.07	µg/mL	0.00052	0.0010	0.067
11	Control Spot	N/A	N/A	N/A	N/A	N/A
12	REF	N/A	N/A	N/A	N/A	N/A


QC Release:



Date:

11 APR 25

QA Release:



Date:

11 APR 25

for Research Use Only