

101896GR - 8.1

QUANSYS BIOSCIENCES
CERTIFICATE OF ANALYSIS


Quansys Biosciences
 365 N 600 W
 Logan, UT 84321
 (435) 752-0531

Q-Plex™ Human Pneumococcal IgG (14-Plex)

REF	Part Number: 482249HU	Software Product Code: HPNM230726-14.3 
LOT	Lot Number: HPNK230726 22124 001	
	Expiration Date: 2026-JAN-07	

Reconstitute Calibrator using 500 µL of Sample Diluent

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	482252HU	HPNP250312	N/A	2026-MAR-12
Sample Diluent	481157HU	HPNS250107	N/A	2026-JAN-07
Calibrator	481155HU	HPNM230726	500 µL	2026-APR-10
Detection	482156HU	HPND250121	N/A	2026-JAN-21
Substrate A	101159GR	QNSA250227	N/A	2026-DEC-09
Substrate B+	101120GR	QNSB+250227	N/A	2026-DEC-09
Wash Buffer	101158GR	QNSW250106	N/A	2027-JAN-06



Q-Plex™ ARRAYS

Spot #	Analyte	Cal Conc	Units	LLD	LLOQ	ULOQ
1	1	0.85	µg/mL	0.00042	0.00083	0.21
2	3	0.15	µg/mL	0.00028	0.00057	0.15
3	4	0.33	µg/mL	0.00016	0.00033	0.083
4	5	0.75	µg/mL	0.00037	0.00073	0.19
5	6B	0.91	µg/mL	0.00044	0.00088	0.23
6	7F	0.83	µg/mL	0.0016	0.0032	0.21
7	8	1.42	µg/mL	0.0028	0.0056	0.36
8	9N	0.70	µg/mL	0.0014	0.0027	0.18
9	9V	0.64	µg/mL	0.0013	0.0025	0.16
10	12F	0.22	µg/mL	0.00043	0.00086	0.055
11	14	3.80	µg/mL	0.0019	0.0037	0.24
12	18C	0.73	µg/mL	0.0014	0.0029	0.18
13	19F	1.46	µg/mL	0.0029	0.0057	0.37
14	23F	0.60	µg/mL	0.0012	0.0023	0.15
15	CWPS	N/A	N/A	N/A	N/A	N/A
16	Control Spot	N/A	N/A	N/A	N/A	N/A
17	REF	N/A	N/A	N/A	N/A	N/A

QC Release:

QA Release:

Date:

Date:

for Research Use Only