

101896GR - 8.1

QUANSYS BIOSCIENCES
CERTIFICATE OF ANALYSIS


Quansys Biosciences
365 N 600 W
Logan, UT 84321
(435) 752-0531

Q-Plex™ Human Pneumococcal IgG (9-Plex)

REF	Part Number: 482149HU-30-25	Software Product Code: HPNM250409-9.3 
LOT	Lot Number: HPNK250409 22475 001	
	Expiration Date: 2025-DEC-19	

Reconstitute Calibrator using 500 µL of Sample Diluent

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	482152HU	HPNP250505	N/A	2026-MAY-05
Sample Diluent	481157HU-30	HPNS241219	N/A	2025-DEC-19
Calibrator	481155HU	HPNM250409	500 µL	2026-JUL-09
Detection	482156HU	HPND250512	N/A	2026-MAY-12
Substrate A	101159GR	QNSA250113	N/A	2026-DEC-09
Substrate B+	101120GR	QNSB+250113	N/A	2026-DEC-09
Wash Buffer	101158GR	QNSW250128	N/A	2027-JAN-28

**Q-Plex™ ARRAYS**

Spot #	Analyte	Cal Conc	Units	LLD	LLOQ	ULOQ
1	CWPS	N/A	N/A	N/A	N/A	N/A
2	2	1.95	µg/mL	0.0016	0.0031	0.43
3	10A	1.05	µg/mL	0.00085	0.0017	0.057
4	11A	0.42	µg/mL	0.00030	0.00060	0.091
5	15B	1.28	µg/mL	0.0011	0.0021	0.29
6	17F	0.63	µg/mL	0.0022	0.0043	0.14
7	19A	1.17	µg/mL	0.0036	0.0073	0.24
8	20	0.88	µg/mL	0.00070	0.0014	0.18
9	22F	0.74	µg/mL	0.0024	0.0048	0.17
10	33F	0.89	µg/mL	0.00070	0.0014	0.047
11	Control Spot	N/A	N/A	N/A	N/A	N/A
12	REF	N/A	N/A	N/A	N/A	N/A

QC Release:

Date:

09 JUL 25

QA Release:

Date:

09 JUL 25

for Research Use Only