

101896GR - 8.1

QUANSYS BIOSCIENCES
CERTIFICATE OF ANALYSIS


Quansys Biosciences
 365 N 600 W
 Logan, UT 84321
 (435) 752-0531

Q-Plex™ Human Cytokine Panel 1 (6-Plex)

REF	Part Number: 115233HU	Software Product Code: HA1M220719-P16P 
LOT	Lot Number: HCP1K220719 22725 001	
	Expiration Date: 2026-MAR-19	

Reconstitute Calibrator using 650 µL of Sample Diluent

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	115252HU	HCP1P250508	N/A	2027-MAY-08
Sample Diluent	101104HU	HQ6S240805	N/A	2026-AUG-05
Calibrator	102655HU	HA1M220719	650 µL	2027-SEP-17
Detection	110956HU 01	HCSDQ6250818	N/A	2027-AUG-18
Streptavidin HRP	101173GR	QNSN241105	N/A	2026-MAY-05
Substrate A	101159GR	QNSA240531	N/A	2026-MAR-19
Substrate B+	101120GR	QNSB+240531	N/A	2026-MAR-19
Wash Buffer	101158GR	QNSW241030	N/A	2026-OCT-30



Spot #	Analyte	Cal Conc	Units	LLD	LLOQ	ULOQ
1	IL-1β	11000	pg/mL	7.54	15.09	11000
2	IL-6	2710	pg/mL	1.86	3.72	2710
3	TNFα	5490	pg/mL	3.77	7.53	5490
4	IFNγ	8520	pg/mL	5.84	11.69	8520
5	IL-1α	3650	pg/mL	2.50	5.01	3650
6	IL-10	7870	pg/mL	5.40	10.80	7870
7	REF	N/A	N/A	N/A	N/A	N/A

Q-Plex™ ARRAYS

QC Release:




Date:

25 SEP 25

QA Release:

Date:

25 SEP 25

for Research Use Only