

101896GR - 8.1

QUANSYS
 BIOSCIENCES
CERTIFICATE OF ANALYSIS


Quansys Biosciences
 365 N 600 W
 Logan, UT 84321
 (435) 752-0531

Q-Plex™ Human Custom 10-Plex

REF	Part Number: 107749GR-10	Software Product Code:
LOT	Lot Number: HCUK220719 22695 001	HA1M220719-L10
	Expiration Date: 2026-JUL-24	

Reconstitute Calibrator 1 using 650 µL of Sample Diluent
Reconstitute Calibrator 2 using 400 µL of Calibrator 1

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	101852GR	HCUP250827	N/A	2027-AUG-27
Sample Diluent	101104HU	HQ6S250205	N/A	2027-FEB-05
Calibrator 1	102655HU	HA1M220719	650 µL	2027-SEP-11
Calibrator 2	102855HU	HA2M200303	400 µL	2027-SEP-11
Detection	101856GR	HCUD250903	N/A	2027-SEP-03
Streptavidin HRP	101173GR	QNSN250124	N/A	2026-JUL-24
Substrate A	101159GR	QNSA250508	N/A	2026-DEC-09
Substrate B+	101120GR	QNSB+250508	N/A	2026-DEC-09
Wash Buffer	101158GR	QNSW241202	N/A	2026-DEC-02



Q-Plex™ ARRAYS

Spot #	Analyte	Cal Conc	Units	LLD	LLOQ	ULOQ
1	IFNα	340	pg/mL	0.23	0.47	340
2	IFNγ	8520	pg/mL	5.84	11.69	8520
3	IFNλ	12480	pg/mL	8.56	17.12	12480
4	IL-1β	11000	pg/mL	7.54	15.09	11000
5	IL-2	4650	pg/mL	3.19	6.38	4650
6	IL-6	2710	pg/mL	1.86	3.72	2710
7	IL-8	1880	pg/mL	1.29	2.58	1880
8	IL-10	7870	pg/mL	5.40	10.80	7870
9	IP-10	3240	pg/mL	2.22	4.44	1080
10	TNFα	5490	pg/mL	3.77	7.53	5490
11	REF	N/A	N/A	N/A	N/A	N/A


QC Release:



Date:

11SEP25

QA Release:



Date:

11SEP25

for Research Use Only