

101896GR - 8.1

QUANSYS
 BIOSCIENCES
CERTIFICATE OF ANALYSIS


Quansys Biosciences
 365 N 600 W
 Logan, UT 84321
 (435) 752-0531

Q-Plex™ Human Cytokine Inflammation Panel 1 (9-Plex)

REF	Part Number: 110433HU	Software Product Code: HA1M220719-IF 
LOT	Lot Number: HCIK220719 22900 001	
	Expiration Date: 2026-APR-12	

Reconstitute Calibrator using 650 µL of Sample Diluent

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	110452HU	HCIP240412	N/A	2026-APR-12
Sample Diluent	101104HU	HQ6S240805	N/A	2026-AUG-05
Calibrator	102655HU	HA1M220719	650 µL	2027-OCT-31
Detection	110956HU_01	HCSDQ6250818	N/A	2027-AUG-18
Streptavidin HRP	101173GR	QNSN250124	N/A	2026-JUL-24
Substrate A	101159GR	QNSA250508	N/A	2026-DEC-09
Substrate B+	101120GR	QNSB+250508	N/A	2026-DEC-09
Wash Buffer	101158GR	QNSW250812	N/A	2027-AUG-12

**Q-Plex™** ARRAYS

Spot #	Analyte	Cal Conc	Units	LLD	LLOQ	ULOQ
1	IL-1α	3650	pg/mL	2.50	5.01	3650
2	IL-1β	11000	pg/mL	7.54	15.09	11000
3	IL-2	4650	pg/mL	3.19	6.38	4650
4	IL-4	1540	pg/mL	1.06	2.11	1540
5	IL-6	2710	pg/mL	1.86	3.72	2710
6	IL-8	1880	pg/mL	1.29	2.58	1880
7	IL-10	7870	pg/mL	5.40	10.80	7870
8	IFNγ	8520	pg/mL	5.84	11.69	8520
9	TNFα	5490	pg/mL	3.77	7.53	5490
10	REF	N/A	N/A	N/A	N/A	N/A


QC Release:



Date:

3100525

QA Release:



Date:

3100525

for Research Use Only