

101896GR - 8.1

QUANSYS BIOSCIENCES
CERTIFICATE OF ANALYSIS


Quansys Biosciences
365 N 600 W
Logan, UT 84321
(435) 752-0531

Q-Plex™ Human Micronutrient v2 (7-Plex)

REF	Part Number: 355149HU	Software Product Code: HMTMA251021C 
LOT	Lot Number: HMTK250318 22706 001	
	Expiration Date: 2026-DEC-09	

Reconstitute Calibrator using 270 uL of Complete Sample Diluent and 30 uL of Calibrator Additive

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	355152HU	HMTTP250929	N/A	2027-MAR-29
Sample Diluent	355157HU	HMTS251001	N/A	2027-APR-01
Competitor	355190HU	HMTC250407	20 mL	2027-MAY-12
Calibrator	355155HU	HMTM250318	270 uL	2027-MAY-12
Calibrator Additive	355142HU	HMTA251021	N/A	2027-APR-21
Detection	355156HU	HMTD251016	N/A	2027-APR-16
Streptavidin HRP	101173GR	QNSN250610	N/A	2026-DEC-10
Substrate A	101159GR	QNSA251013	N/A	2026-DEC-09
Substrate B2	101128GR	QNSB2251013	N/A	2026-DEC-09
Wash Buffer	101158GR	QNSW251031	N/A	2027-OCT-31



Spot #	Analyte	Cal Conc	Units	LLD	LLOQ	ULOQ
1	Ferritin	25.56	µg/L	0.018	0.035	25.56
2	*AGP	0.38	g/L	0.00078	0.0016	0.38
3	Tg	16.10	µg/L	0.012	0.024	16.10
4	sTfR	4.30	mg/L	0.0029	0.0059	4.30
5	REF	N/A	N/A	N/A	N/A	N/A
6	CRP	0.97	mg/L	0.00067	0.0013	0.17
7	HRP2	1.77	µg/L	0.0051	0.010	1.77
8	*RBP4	0.97	µmol/L	0.0020	0.0040	0.97

Q-Plex™ ARRAYS
*Competitive Assay

QC Release:

Date:

12NOV25

QA Release:

Date:

12NOV25

for Research Use Only