

101896GR - 8.1

QUANSYS BIOSCIENCES
CERTIFICATE OF ANALYSIS


Quansys Biosciences
 365 N 600 W
 Logan, UT 84321
 (435) 752 - 0531

Q-Plex™ Human Pneumococcal IgG (14-Plex) Aliquoted Bulk 30 mL Diluent

REF	Part Number: 482249HU-30-25	Software Product Code: HPNM250409-A14.3 
LOT	Lot Number: HPNK250409 22985 001	
	Expiration Date: 2026-MAY-12	

Reconstitute Calibrator using 500 µL of Sample Diluent

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	482252HU	HPNP251006	N/A	2026-OCT-06
Sample Diluent	481157HU-30	HPNS251028	N/A	2026-OCT-28
Calibrator	481155HU	HPNM250409	500 µL	2026-NOV-13
Detection	482156HU	HPND250512	N/A	2026-MAY-12
Substrate A	101159GR	QNSA250508	N/A	2026-DEC-09
Substrate B+	101120GR	QNSB+250508	N/A	2026-DEC-09
Wash Buffer	101158GR	QNSW251031	N/A	2027-OCT-31



Q-Plex™ ARRAYS

Spot #	Analyte	Cal Conc	Units	LLD	LLOQ	ULOQ
1	1	1.16	µg/mL	0.00095	0.0019	0.26
2	3	0.33	µg/mL	0.0011	0.0021	0.29
3	4	0.69	µg/mL	0.0055	0.0011	0.15
4	5	1.63	µg/mL	0.0014	0.0027	0.37
5	6B	0.54	µg/mL	0.00050	0.0010	0.13
6	7F	1.71	µg/mL	0.0057	0.011	0.39
7	8	0.58	µg/mL	0.0019	0.0038	0.13
8	9N	0.63	µg/mL	0.0022	0.0044	0.15
9	9V	0.92	µg/mL	0.0029	0.0057	0.20
10	12F	0.32	µg/mL	0.0011	0.0021	0.070
11	14	2.06	µg/mL	0.0018	0.0035	0.12
12	18C	0.97	µg/mL	0.0031	0.0062	0.22
13	19F	4.81	µg/mL	0.014	0.027	0.92
14	23F	0.36	µg/mL	0.0013	0.0025	0.084
15	CWPS	N/A	N/A	N/A	N/A	N/A
16	Control Spot	N/A	N/A	N/A	N/A	N/A
17	REF	N/A	N/A	N/A	N/A	N/A

QC Release:

Date: 13 NOV 25

QA Release:

for Research Use Only

Date: 13 Nov 25