

101896GR - 8.1

QUANSYS BIOSCIENCES
CERTIFICATE OF ANALYSIS


Quansys Biosciences
 365 N 600 W
 Logan, UT 84321
 (435) 752-0531

Q-Plex™ Human Cytokine Panel 2 (4-Plex)

REF	Part Number: 112033HU	Software Product Code: HA1M190516P2Q 
LOT	Lot Number: HCP2K190516 23115 001	
	Expiration Date: 2026-APR-10	

Reconstitute Calibrator using 600 µL of Sample Diluent

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	112052HU	HCP2P240410	N/A	2026-APR-10
Sample Diluent	101104HU	HQ6S250205	N/A	2027-FEB-05
Calibrator	102655HU	HA1M190516	600 µL	2027-DEC-11
Detection	110956HU 01	HCSDQ6250818	N/A	2027-AUG-18
Streptavidin HRP	101173GR	QNSN250124	N/A	2026-JUL-24
Substrate A	101159GR	QNSA251013	N/A	2026-DEC-09
Substrate B+	101120GR	QNSB+251013	N/A	2026-DEC-09
Wash Buffer	101158GR	QNSW250624	N/A	2027-JUN-24



Spot #	Analyte	Cal Conc	Units	LLD	LLOQ	ULOQ
1	IL-1α	3400	pg/mL	2.33	4.66	3400
2	IL-4	1257	pg/mL	0.86	1.72	1257
3	IL-6	2700	pg/mL	1.85	3.70	2700
4	TNFα	4912	pg/mL	3.37	6.74	4912
5	REF	N/A	N/A	N/A	N/A	N/A

Q-Plex™ ARRAYS

QC Release:

Date:

11 DEC 25

QA Release:

Date:

12 DEC 25

for Research Use Only