

101896GR - 8.1

QUANSYS BIOSCIENCES
CERTIFICATE OF ANALYSIS


Quansys Biosciences
 365 N 600 W
 Logan, UT 84321
 (435) 752-0531

Q-Plex™ Human Pneumococcal IgG (14-Plex)

REF	Part Number: 482249HU	Software Product Code:
LOT	Lot Number: HPNK240514 23040 001	HPNM240514-14.3W
	Expiration Date: 2026-OCT-06	

Reconstitute Calibrator using 500 µL of Sample Diluent

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	482252HU	HPNP251006	N/A	2026-OCT-06
Sample Diluent	481157HU	HPNS251028	N/A	2026-OCT-28
Calibrator	481155HU	HPNM240514	500 µL	2026-NOV-24
Detection	482156HU	HPND251111	N/A	2026-NOV-11
Substrate A	101159GR	QNSA251013	N/A	2026-DEC-09
Substrate B+	101120GR	QNSB+251013	N/A	2026-DEC-09
Wash Buffer	101158GR	QNSW250128	N/A	2027-JAN-28



Q-Plex™ ARRAYS

Spot #	Analyte	Cal Conc	Units	LLD	LLOQ	ULOQ
1	1	0.76	µg/mL	0.00037	0.00074	0.19
2	3	0.12	µg/mL	0.00023	0.00047	0.12
3	4	0.27	µg/mL	0.00013	0.00026	0.068
4	5	0.65	µg/mL	0.00032	0.00063	0.16
5	6B	0.73	µg/mL	0.00036	0.00071	0.18
6	7F	0.73	µg/mL	0.0014	0.0029	0.18
7	8	1.27	µg/mL	0.0025	0.0050	0.32
8	9N	0.59	µg/mL	0.0012	0.0023	0.15
9	9V	0.58	µg/mL	0.0011	0.0023	0.15
10	12F	0.20	µg/mL	0.00039	0.00078	0.050
11	14	3.49	µg/mL	0.0017	0.0034	0.22
12	18C	0.65	µg/mL	0.0013	0.0025	0.16
13	19F	1.27	µg/mL	0.0025	0.0050	0.32
14	23F	0.57	µg/mL	0.0011	0.0022	0.14
15	CWPS	N/A	N/A	N/A	N/A	N/A
16	Control Spot	N/A	N/A	N/A	N/A	N/A
17	REF	N/A	N/A	N/A	N/A	N/A

QC Release:

Kayla Kowalik

Date:

24 Nov 25

QA Release:

Stacy Hange
 for Research Use Only

Date:

24 Nov 25