

101896GR - 8.1

**QUANSYS** BIOSCIENCES  
**CERTIFICATE OF ANALYSIS**


Quansys Biosciences  
 365 N 600 W  
 Logan, UT 84321  
 (435) 752-0531

**Q-Plex™ Human Pneumococcal IgG (9-Plex)**

<b>REF</b>	Part Number: 482149HU	Software Product Code: <b>HPNM240514-9.3</b> 
<b>LOT</b>	Lot Number: HPNK240514 23211 001	
	Expiration Date: 2026-NOV-03	

**Reconstitute Calibrator using 500 µL of Sample Diluent**

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	482152HU	HPNP251103	N/A	2026-NOV-03
Sample Diluent	481157HU	HPNS251204	N/A	2026-DEC-04
Calibrator	481155HU	HPNM240514	500 µL	2027-JAN-23
Detection	482156HU	HPND260119	N/A	2027-JAN-19
Substrate A	101159GR	QNSA251218	N/A	2026-DEC-09
Substrate B+	101120GR	QNSB+251218	N/A	2026-DEC-09
Wash Buffer	101158GR	QNSW260105	N/A	2028-JAN-05



Spot #	Analyte	Cal Conc	Units	LLD	LLOQ	ULOQ
1	CWPS	N/A	N/A	N/A	N/A	N/A
2	2	2.19	µg/mL	0.0016	0.0031	0.43
3	10A	1.16	µg/mL	0.00085	0.0017	0.057
4	11A	0.44	µg/mL	0.00030	0.00060	0.091
5	15B	1.51	µg/mL	0.0011	0.0021	0.29
6	17F	0.74	µg/mL	0.0022	0.0043	0.14
7	19A	1.36	µg/mL	0.0036	0.0073	0.24
8	20	0.91	µg/mL	0.00070	0.0014	0.18
9	22F	0.85	µg/mL	0.0024	0.0048	0.17
10	33F	0.96	µg/mL	0.00070	0.0014	0.047
11	Control Spot	N/A	N/A	N/A	N/A	N/A
12	REF	N/A	N/A	N/A	N/A	N/A

**Q-Plex™ ARRAYS**

QC Release: *Kayla Kowallis* Date: 23 JAN 26  
 QA Release: *[Signature]* Date: 23 JAN 26  
 for Research Use Only