



CERTIFICATE OF ANALYSIS


Q-Plex™ Human Pneumococcal IgG (9-Plex) 30 mL Diluent

REF	Part Number: 482149HU-30	Software Product Code: HPNM250409-9.3 
LOT	Lot Number: HPNK250409 23366 001	
	Expiration Date: 2026-NOV-11	

Reconstitute Calibrator using 500 µL of Sample Diluent

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	482152HU	HPNP260127	N/A	2027-JAN-27
Sample Diluent	481157HU-30	HPNS251204	N/A	2026-DEC-04
Calibrator	481155HU	HPNM250409	500 µL	2027-FEB-11
Detection	482156HU	HPND251111	N/A	2026-NOV-11
Substrate A	101159GR	QNSA251218	N/A	2026-DEC-09
Substrate B+	101120GR	QNSB+251218	N/A	2026-DEC-09
Wash Buffer	101158GR	QNSW260123	N/A	2028-JAN-23

Spot #	Analyte	Cal Conc	Units	LLD	LLOQ	ULOQ
1	CWPS	N/A	N/A	N/A	N/A	N/A
2	2	1.95	µg/mL	0.0016	0.0031	0.43
3	10A	1.05	µg/mL	0.00085	0.0017	0.057
4	11A	0.42	µg/mL	0.00030	0.00060	0.091
5	15B	1.28	µg/mL	0.0011	0.0021	0.29
6	17F	0.63	µg/mL	0.0022	0.0043	0.14
7	19A	1.17	µg/mL	0.0036	0.0073	0.24
8	20	0.88	µg/mL	0.00070	0.0014	0.18
9	22F	0.74	µg/mL	0.0024	0.0048	0.17
10	33F	0.89	µg/mL	0.00070	0.0014	0.047
11	Control Spot	N/A	N/A	N/A	N/A	N/A
12	REF	N/A	N/A	N/A	N/A	N/A



QC Release: *Kayla Kowalkis* Date: 11 FEB 20

QA Release: *[Signature]* Date: 11 FEB 20