


# Q-Plex™ ARRAY


## Q-Plex™ SARS-CoV-2 Human IgG (5-Plex)

Software Product Code: HCVM200715 19534 001  
 S1 Cutoff Value: 7.7  
 S2 Cutoff Value: 30  
 NP Cutoff Value: 25

<b>REF</b>	721649HU
<b>LOT</b>	HCVK200715 23538 001
	2025-JUL-20

Product Information				
Component	Part Number	Lot Number	Exp. Date	Recon. Volume
Plate	721652HU	HCVNP240110	2026-JAN-10	N/A
Sample Diluent	691657HU	HCVS251014	2026-OCT-14	N/A
Calibrator	691655HU	HCVM200715	2028-MAR-12	300 µL
High Control	691631HU	HCVM200715-H	2028-MAR-12	300 µL
Low Control	691632HU	HCVM200715-L	2028-MAR-12	300 µL
Detection	691656HU	HCVD230720	2025-JUL-20	N/A
Streptavidin HRP	691173GR	QNSN250610	2026-DEC-10	N/A
Substrate A	691159GR	QNSA250725	2026-DEC-10	N/A
Substrate B+	691120GR	QNSB+250725	2026-DEC-09	N/A
Wash Buffer	691158GR	QNSW250812	2027-AUG-12	N/A

Assay Performance			
Analyte	Calibrator Conc. (U/mL)	High Control Range (U/mL)	Low Control Range (U/mL)
S1	1024	56.96 - 116.59	3.22 - 7.49
S2	910	160.30 - 315.90	30.54 - 51.31
NP	1000	39.68 - 92.60	16.96 - 39.58



Quality Control: *[Signature]* Date: 12 MAR 26

Quality Assurance: *Kanika Kowallio* Date: 12 MAR 26