



CERTIFICATE OF ANALYSIS


Q-Plex™ Human Angiogenesis (9-Plex)

REF	Part Number: 150233HU	Software Product Code: HA1M251204-AG 
LOT	Lot Number: HAGK251204 23719 001	
	Expiration Date: 2026-DEC-09	

Reconstitute Calibrator using 650µL of Complete Sample Diluent

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	150252HU	HAGP260323	N/A	2028-MAR-23
Sample Diluent	101104HU	HQ6S251021	N/A	2027-OCT-21
Calibrator	102655HU	HA1M251204	650µL	2028-APR-23
Detection	150256HU 01	HAGDQ6260106	N/A	2028-JAN-06
Streptavidin HRP	101173GR	QNSN251216	N/A	2027-JUN-16
Substrate A	101159GR	QNSA260226	N/A	2026-DEC-09
Substrate B+	101120GR	QNSB+260226	N/A	2026-DEC-09
Wash Buffer	101158GR	QNSW260310	N/A	2028-MAR-10

Spot #	Analyte	Cal Conc	Units	LLD	LLOQ	ULOQ
1	ANG-2	9410	pg/mL	6.45	12.91	9410
2	FGF	4690	pg/mL	3.22	6.43	4690
3	HGF	12890	pg/mL	8.84	17.68	12890
4	IL-8	2280	pg/mL	1.56	3.13	2280
5	PDGF	4440	pg/mL	3.05	6.09	4440
6	TIMP-1	13540	pg/mL	27.86	55.72	13540
7	TIMP-2	19150	pg/mL	13.13	26.27	19150
8	TNFα	7100	pg/mL	4.87	9.74	7100
9	VEGF	1980	pg/mL	1.36	2.72	1980
10	REF	N/A	N/A	N/A	N/A	N/A



QC Release: *[Signature]* Date: 23 APR 2026

QA Release: *Kayle Kowallis* Date: 23 APR 2026

101896GR - 9.0

for Research Use Only