



**CERTIFICATE OF ANALYSIS**


**Q-Plex™ Human Pneumococcal IgG (9-Plex)**

<b>REF</b>	Part Number: 482149HU	Software Product Code: <b>HPNM260113-9.3</b> 
<b>LOT</b>	Lot Number: HPNK260113 23718 001	
	Expiration Date: 2026-DEC-04	

**Reconstitute Calibrator using 500 µL of Sample Diluent**

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	482152HU	HPNP260127	N/A	2027-JAN-27
Sample Diluent	481157HU	HPNS251204	N/A	2026-DEC-04
Calibrator	481155HU	HPNM260113	500 µL	2027-MAY-07
Detection	482156HU	HPND260119	N/A	2027-JAN-19
Substrate A	101159GR	QNSA260226	N/A	2026-DEC-09
Substrate B+	101120GR	QNSB+260226	N/A	2026-DEC-09
Wash Buffer	101158GR	QNSW260411	N/A	2028-APR-11

Spot #	Analyte	Cal Conc	Units	LLD	LLOQ	ULOQ
1	CWPS	N/A	N/A	N/A	N/A	N/A
2	2	2.46	µg/mL	0.0016	0.0031	0.43
3	10A	1.38	µg/mL	0.00085	0.0017	0.057
4	11A	0.47	µg/mL	0.00030	0.00060	0.091
5	15B	1.53	µg/mL	0.0011	0.0021	0.29
6	17F	0.80	µg/mL	0.0022	0.0043	0.14
7	19A	1.33	µg/mL	0.0036	0.0073	0.24
8	20	0.94	µg/mL	0.00070	0.0014	0.18
9	22F	0.77	µg/mL	0.0024	0.0048	0.17
10	33F	0.98	µg/mL	0.00070	0.0014	0.047
11	Control Spot	N/A	N/A	N/A	N/A	N/A
12	REF	N/A	N/A	N/A	N/A	N/A



QC Release: Kayla Kowallis Date: 07MAY26

QA Release: JL Cox Date: 07MAY26