



CERTIFICATE OF ANALYSIS


Q-Plex™ Human Pneumococcal IgG (9-Plex) 30 mL Diluent

REF	Part Number: 482149HU-30	Software Product Code: HPNM260113-9.3 
LOT	Lot Number: HPNK260113 23764 001	
	Expiration Date: 2026-DEC-09	

Reconstitute Calibrator using 500 µL of Sample Diluent

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	482152HU	HPNP260127	N/A	2027-JAN-27
Sample Diluent	481157HU-30	HPNS260414	N/A	2027-APR-14
Calibrator	481155HU	HPNM260113	500 µL	2027-MAY-26
Detection	482156HU	HPND260119	N/A	2027-JAN-19
Substrate A	101159GR	QNSA260226	N/A	2026-DEC-09
Substrate B+	101120GR	QNSB+260226	N/A	2026-DEC-09
Wash Buffer	101158GR	QNSW260518	N/A	2028-MAY-18

Spot #	Analyte	Cal Conc	Units	LLD	LLOQ	ULOQ
1	CWPS	N/A	N/A	N/A	N/A	N/A
2	2	2.46	µg/mL	0.0016	0.0031	0.43
3	10A	1.38	µg/mL	0.00085	0.0017	0.057
4	11A	0.47	µg/mL	0.00030	0.00060	0.091
5	15B	1.53	µg/mL	0.0011	0.0021	0.29
6	17F	0.80	µg/mL	0.0022	0.0043	0.14
7	19A	1.33	µg/mL	0.0036	0.0073	0.24
8	20	0.94	µg/mL	0.00070	0.0014	0.18
9	22F	0.77	µg/mL	0.0024	0.0048	0.17
10	33F	0.98	µg/mL	0.00070	0.0014	0.047
11	Control Spot	N/A	N/A	N/A	N/A	N/A
12	REF	N/A	N/A	N/A	N/A	N/A



QC Release:

Kaylee Kowallis

Date:

26 MAY 26

QA Release:

[Signature]

Date:

26 MAY 26