



Quansys Biosciences, Inc. | 365 N 600 W Logan, UT 84321 | (435) 752-0531

CERTIFICATE OF ANALYSIS


Q-Plex™ Human Malaria (5-Plex)

REF	Part Number: 565949HU	Software Product Code: HMRM191125-ML2
LOT	Lot Number: HMRK191125 23855 001	
><	Expiration Date: 2026-OCT-14	

Reconstitute Calibrator using 350 µL of Calibrator Diluent

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	565952HU	HMRP260326	N/A	2027-SEP-26
Calibrator Diluent	565928HU	HMRU260515	6 mL	2027-NOV-15
Assay Diluent	565983HU	HMRV260514	N/A	2027-NOV-14
Competitor	565990HU	HMRC191014	6 mL	2027-DEC-04
Calibrator	565955HU	HMRM191125	350 µL	2027-DEC-04
Detection	565956HU	HMRD250414	N/A	2026-OCT-14
Streptavidin HRP	101173GR	QNSN251216	N/A	2027-JUN-16
Substrate A+	105959GR	MOSA+260304	N/A	2029-FEB-28
Substrate B	105920GR	MOSB260304	N/A	2029-FEB-28
Wash Buffer	101158GR	QNSW260518	N/A	2028-MAY-18

Spot #	Analyte	Cal Conc	Units	LLD	LLOQ	ULOQ
1	HRP2	2800	pg/mL	0.53	0.68	2800
2	LDH-Pan	67000	pg/mL	7.28	16.36	67000
3	LDH-Pv	19200	pg/mL	4.23	4.69	19200
4	LDH-Pf	20800	pg/mL	2.33	5.08	20800
5	*CRP	38000	ng/mL	7.65	9.28	38000
6	REF	N/A	N/A	N/A	N/A	N/A



*Competitive Assay

QC Release: Kayla Kowallis Date: 04 JUN 26

QA Release: R Cox Date: 05 JUN 26

101896GR - 9.0

for Research Use Only