



Quansys Biosciences, Inc. | 365 N 600 W Logan, UT 84321 | (435) 752-0531

CERTIFICATE OF ANALYSIS


Q-Plex™ Human Custom 3-Plex

REF	Part Number: 107749GR-03	Software Product Code: HA1M251204-L3P 
LOT	Lot Number: HCUK251204 24067 001	
	Expiration Date: 2027-OCT-21	

Reconstitute Calibrator using 650µL of Sample Diluent

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	101852GR	HCUP260622	N/A	2028-JUN-22
Sample Diluent	101104HU	HQ6S251021	N/A	2027-OCT-21
Calibrator	102655HU	HA1M251204	650µL	2028-JUN-30
Detection	110956HU 01	HCSDQ6260220	N/A	2028-FEB-20
Streptavidin HRP	101173GR	QNSN260507	N/A	2027-NOV-07
Substrate A	101159GR	QNSA260427	N/A	2028-APR-02
Substrate B+	101120GR	QNSB+260427	N/A	2028-APR-02
Wash Buffer	101158GR	QNSW260518	N/A	2028-MAY-18

Spot #	Analyte	Cal Conc	Units	LLD	LLOQ	ULOQ
1	IL-6	3210	pg/mL	2.20	4.40	3210
2	IL-8	2280	pg/mL	1.56	3.13	2280
3	IL-10	9050	pg/mL	6.21	12.41	9050
4	REF	N/A	N/A	N/A	N/A	N/A



Q-Plex™ ARRAYS

QC Release: Kayla Kowallis Date: 30 JUN 26

QA Release: R Cox Date: 30 JUN 26

101896GR - 9.0

For Research Use Only