

Q-Plex™ ARRAY


Q-Plex SARS-CoV-2 Human IgG (5-Plex)

Software Product Code: HCV200715 23881 001
 S1 Cutoff Value: 7.7 U/mL
 S2 Cutoff Value: 30 U/mL

REF	721649HU
LOT	HCV200715 23881 001
	2027-JUN-15

Product Information				
Component	Part Number	Lot Number	Exp. Date	Recon. Volume
Plate	721652HU	HCVNP260601	2028-JUN-01	N/A
Sample Diluent	691657HU	HCSV260615	2027-JUN-15	N/A
Calibrator	691655HU	HCV200715	2028-JUL-07	300 µL
High Control	691631HU	HCV200715-H	2028-JUL-07	300 µL
Low Control	691632HU	HCV200715-L	2028-JUL-07	300 µL
Detection	691656HU	HCVD260603	2029-JUN-03	N/A
Streptavidin HRP	101173GR	QNSN260507	2027-NOV-07	N/A
Substrate A	101159GR	QNSA260427	2028-APR-02	N/A
Substrate B+	101120GR	QNSB+260427	2028-APR-02	N/A
Wash Buffer	101158GR	QNSW260227	2028-FEB-27	N/A

Assay Performance			
Analyte	Calibrator Conc. (U/mL)	High Control Range (U/mL)	Low Control Range (U/mL)
S1	700	56.96 - 116.59	3.22 - 7.49
S2	780	160.30 - 315.90	30.54 - 51.31



Quality Control: R Cux Date: 07JUL26

Quality Assurance: [Signature] Date: 07JUL26