



Quansys Biosciences, Inc. | 365 N 600 W Logan, UT 84321 | (435) 752-0531

CERTIFICATE OF ANALYSIS


Q-Plex™ Human Pneumococcal IgG (9-Plex) Aliquoted Bulk 30 mL Diluent

REF	Part Number: 482149HU-30-25	Software Product Code:
LOT	Lot Number: HPNK260514 24047 001	HPNM260514 23763 001
EXP	Expiration Date: 2027-MAY-08	

Reconstitute Calibrator using 500µL of Sample Diluent

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	482152HU	HPNP260527	N/A	2027-MAY-27
Sample Diluent	481157HU-30	HPNS260624	N/A	2027-JUN-24
Calibrator	481155HU	HPNM260514	500µL	2027-JUL-15
Detection	482156HU	HPND260508	N/A	2027-MAY-08
Substrate A	101159GR	QNSA260427	N/A	2028-APR-02
Substrate B+	101120GR	QNSB+260427	N/A	2028-APR-02
Wash Buffer	101158GR	QNSW260604	N/A	2028-JUN-04

Spot #	Analyte	Cal Conc	Units	LLD	LLOQ	ULOQ
1	CWPS	N/A	N/A	N/A	N/A	N/A
2	2	2.45	µg/mL	0.00063	0.0013	1.23
3	10A	1.28	µg/mL	0.00044	0.00088	0.45
4	11A	0.46	µg/mL	0.00014	0.00029	0.14
5	15B	1.63	µg/mL	0.00058	0.0012	0.60
6	17F	0.85	µg/mL	0.00044	0.00089	0.52
7	19A	1.42	µg/mL	0.00065	0.0013	1.31
8	20	0.99	µg/mL	0.00034	0.00068	0.59
9	22F	1.06	µg/mL	0.00042	0.00083	0.64
10	33F	1.08	µg/mL	0.00052	0.0010	0.35
11	Control Spot	N/A	N/A	N/A	N/A	N/A
12	REF	N/A	N/A	N/A	N/A	N/A



Q-Plex™ ARRAYS

QC Release:

Kaylee Kowallis

Date:

15 JUL 20

QA Release:

Anthony Nails

Date:

15 JUL 20

For Research Use Only

101896GR - 9.0