



Quansys Biosciences, Inc. | 365 N 600 W Logan, UT 84321 | (435) 752-0531

CERTIFICATE OF ANALYSIS


Q-Plex™ Human Pneumococcal IgG (9-Plex)

REF	Part Number: 482149HU	Software Product Code: HPNM260528-9.3 
LOT	Lot Number: HPNK260528 24147 001	
	Expiration Date: 2026-NOV-03	

Reconstitute Calibrator using 500µL of Sample Diluent

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	482152HU	HPNP251103	N/A	2026-NOV-03
Sample Diluent	481157HU	HPNS260624	N/A	2027-JUN-24
Calibrator	482155HU	HPNM260528	500µL	2027-JUL-21
Detection	482156HU	HPND260716	N/A	2027-JUL-16
Substrate A	101159GR	QNSA260427	N/A	2028-APR-02
Substrate B+	101120GR	QNSB+260427	N/A	2028-APR-02
Wash Buffer	101158GR	QNSW260707	N/A	2028-JUL-07

Spot #	Analyte	Cal Conc	Units	LLD	LLOQ	ULOQ
1	CWPS	N/A	N/A	N/A	N/A	N/A
2	2	1.13	µg/mL	0.00053	0.0011	1.13
3	10A	1.34	µg/mL	0.00036	0.00072	0.34
4	11A	0.31	µg/mL	0.000038	0.000076	0.071
5	15B	1.73	µg/mL	0.00023	0.00047	1.73
6	17F	0.40	µg/mL	0.00080	0.0016	0.25
7	19A	0.85	µg/mL	0.00042	0.00083	0.85
8	20	1.17	µg/mL	0.00015	0.00030	0.19
9	22F	1.50	µg/mL	0.00018	0.00037	0.36
10	33F	1.35	µg/mL	0.00016	0.00033	0.64
11	Control Spot	N/A	N/A	N/A	N/A	N/A
12	REF	N/A	N/A	N/A	N/A	N/A



Q-Plex™ ARRAYS

QC Release:

Haylee Kowallis

Date:

21 JUL 26

QA Release:

Anthony Nardiz

Date:

21 JUL 24

101896GR - 9.0

For Research Use Only