



Quansys Biosciences, Inc. | 365 N 600 W Logan, UT 84321 | (435) 752-0531

CERTIFICATE OF ANALYSIS


Q-Plex™ Human Pneumococcal IgG (14-Plex)

REF	Part Number: 482249HU	Software Product Code: HPNM260528-W14.3 
LOT	Lot Number: HPNK260528 24093 001	
EXP	Expiration Date: 2027-MAY-11	

Reconstitute Calibrator using 500µL of Sample Diluent

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	482252HU	HPNP260511	N/A	2027-MAY-11
Sample Diluent	481157HU	HPNS260624	N/A	2027-JUN-24
Calibrator	482155HU	HPNM260528	500µL	2027-JUL-31
Detection	482156HU	HPND260716	N/A	2027-JUL-16
Substrate A	101159GR	QNSA260612	N/A	2028-APR-02
Substrate B+	101120GR	QNSB+260612	N/A	2028-APR-02
Wash Buffer	101158GR	QNSW260604	N/A	2028-JUN-04

Spot #	Analyte	Cal Conc	Units	LLD	LLOQ	ULOQ
1	1	0.50	µg/mL	0.000061	0.00012	0.13
2	3	0.099	µg/mL	0.00012	0.00024	0.097
3	4	0.15	µg/mL	0.000019	0.000037	0.15
4	5	0.14	µg/mL	0.000063	0.00013	0.14
5	6B	0.59	µg/mL	0.000080	0.00016	0.56
6	7F	0.50	µg/mL	0.000061	0.00012	0.32
7	8	0.51	µg/mL	0.00014	0.00028	0.51
8	9N	0.58	µg/mL	0.000078	0.00016	0.33
9	9V	0.59	µg/mL	0.000073	0.00015	0.36
10	12F	0.20	µg/mL	0.000056	0.00011	0.19
11	14	0.72	µg/mL	0.00033	0.00067	0.72
12	18C	0.23	µg/mL	0.000030	0.000059	0.057
13	19F	1.01	µg/mL	0.00028	0.00056	1.01
14	23F	0.77	µg/mL	0.00010	0.00020	0.68
15	CWPS	N/A	N/A	N/A	N/A	N/A
16	Control Spot	N/A	N/A	N/A	N/A	N/A
17	REF	N/A	N/A	N/A	N/A	N/A



Q-Plex™ ARRAYS

QC Release: Kyle Howalls

Date: 31 JUL 26

QA Release: Anthony Nault

Date: 31 JUL 26

101896GR - 9.0

For Research Use Only