QUANSYS BIOSCIENCES

Quansys Biosciences, Inc. | 365 N 600 W Logan, UT 84321 | (435) 752-0531

CERTIFICATE OF ANALYSIS

Q-Plex™ Human Custom 18-Plex

REF	Part Number: 676351HU	Software Product Code: HA1M251204 24146 001
LOT	Lot Number: HCUK251204 24356 001	
	Expiration Date: 2027-SEP-17	

Reconstitute Calibrator using 650 µL of Sample Diluent

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	676352HU	HCUP260917	N/A	2027-SEP-17
Sample Diluent	110957HU	HCSS260820	N/A	2028-JUL-20
Calibrator	102655HU	HA1M251204	650 µL	2027-SEP-29
Detection	676356HU	HCUD260921	N/A	2028-AUG-06
Streptavidin HRP	101173GR	QNSN260507	N/A	2027-NOV-07
Substrate A	101159GR	QNSA260612	N/A	2028-APR-02
Substrate B+	101120GR	QNSB+260612	N/A	2028-APR-02
Wash Buffer	101158GR	QNSW260707	N/A	2028-JUL-07

Spot #	Analyte	Cal Conc	Units	LLD	LLOQ	ULOQ
1	TNFRI	3460	pg/mL	2.37	4.75	3460
2	IL-1α	4540	pg/mL	3.11	6.23	4540
3	IL-1ß	11030	pg/mL	7.57	15.13	11030
4	IL-2	4110	pg/mL	2.82	5.64	4110
5	IL-4	2110	pg/mL	1.45	2.89	2110
6	IL-5	3090	pg/mL	2.12	4.24	3090
7	IL-6	3210	pg/mL	2.20	4.40	3210
8	IL-8	2280	pg/mL	1.56	3.13	2280
9	IFNy	8930	pg/mL	6.12	12.25	8930
10	IL-12p70	3000	pg/mL	2.06	4.12	3000
11	IL-13	3460	pg/mL	2.37	4.75	3460
12	IL-23	37300	pg/mL	25.58	51.17	37300
13	IL-15	3010	pg/mL	2.06	4.13	3010
14	IL-17A	4300	pg/mL	2.95	5.90	4300
15	IL-10	3380	pg/mL	2.32	4.64	3380
16	TNFα	4770	pg/mL	3.27	6.54	4770
17	TNFRII	2010	pg/mL	1.38	2.76	2010
18	TNFß	4530	pg/mL	3.11	6.21	4530
19	REF	N/A	N/A	N/A	N/A	N/A



Q-Plex™ ARRAYS

QC Release: Krzysztof Kowalski Date: 29SEP26

QA Release: Date: 29SEP26

For Research Use Only