QUANSYS BIOSCIENCES

Quansys Biosciences, Inc. | 365 N 600 W Logan, UT 84321 | (435) 752-0531

CERTIFICATE OF ANALYSIS

Q-Plex™ Human Pneumococcal IgG (9-Plex)

REF	Part Number: 482149HU	Software Product Code: **HPNM260528-9.3**
LOT	Lot Number: HPNK260528 24334 001	
	Expiration Date: 2027-MAY-27	

Reconstitute Calibrator using 500 µL of Sample Diluent

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	482152HU	HPNP260527	N/A	2027-MAY-27
Sample Diluent	481157HU	HPNS260818	N/A	2027-AUG-18
Calibrator	482155HU	HPNM260528	500 µL	2027-SEP-08
Detection	482156HU	HPND260811	N/A	2027-AUG-11
Substrate A	101159GR	QNSA260612	N/A	2028-APR-02
Substrate B+	101120GR	QNSB+260612	N/A	2028-APR-02
Wash Buffer	101158GR	QNSW260707	N/A	2028-JUL-07

Spot #	Analyte	Cal Conc	Units	LLD	LLOQ	ULOQ
1	CWPS	N/A	N/A	N/A	N/A	N/A
2	2	1.13	µg/mL	0.00053	0.0011	1.13
3	10A	1.34	µg/mL	0.00036	0.00072	0.34
4	11A	0.31	µg/mL	0.000038	0.000076	0.071
5	15B	1.73	µg/mL	0.00023	0.00047	1.73
6	17F	0.40	µg/mL	0.00080	0.0016	0.25
7	19A	0.85	µg/mL	0.00042	0.00083	0.85
8	20	1.17	µg/mL	0.00015	0.00030	0.19
9	22F	1.50	µg/mL	0.00018	0.00037	0.36
10	33F	1.35	µg/mL	0.00016	0.00033	0.64
11	Control Spot	N/A	N/A	N/A	N/A	N/A
12	REF	N/A	N/A	N/A	N/A	N/A



Q-Plex™ ARRAYS

QC Release: Courtney Nardin Date: 08SEP26

QA Release: Kayla Kowalis Date: 08SEP 26

For Research Use Only