QUANSYS BIOSCIENCES
Quansys Biosciences, Inc. | 365 N 600 W Logan, UT 84321 | (435) 752-0531

CERTIFICATE OF ANALYSIS

Q-Plex™ Human Custom 7-Plex

REF	Part Number: 107749GR-07	Software Product Code: HPNM260528 24404 001
LOT	Lot Number: HCUK260528 24404 001	
	Expiration Date: 2027-MAY-08	

Reconstitute Calibrator using 500 µL of Sample Diluent

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	101852GR	HCUP260925	N/A	2027-SEP-25
Sample Diluent	481157HU	HPNS260928	N/A	2027-SEP-28
Calibrator	482155HU	HPNM260528	500 µL	2027-SEP-30
Detection	482156HU	HPND260508	N/A	2027-MAY-08
Substrate A	101159GR	QNSA260612	N/A	2028-APR-02
Substrate B+	101120GR	QNSB+260612	N/A	2028-APR-02
Wash Buffer	101158GR	QNSW260707	N/A	2028-JUL-07

Spot #	Analyte	Cal Conc	Units	LLD	LLOQ	ULOQ
1	6B	0.59	µg/mL	0.00018	0.00035	0.59
2	10A	1.34	µg/mL	0.00035	0.00071	1.32
3	14	0.72	µg/mL	0.00021	0.00042	0.71
4	15B	1.73	µg/mL	0.00022	0.00043	1.11
5	17F	0.40	µg/mL	0.00043	0.00085	0.40
6	19A	0.85	µg/mL	0.00040	0.00080	0.82
7	22F	1.50	µg/mL	0.00020	0.00040	0.93
8	CWPS	N/A	N/A	N/A	N/A	N/A
9	Control Spot	N/A	N/A	N/A	N/A	N/A
10	REF	N/A	N/A	N/A	N/A	N/A



Q-Plex™ ARRAYS

QC Release: Krystle Kowalski Date: 30SEP26

QA Release: Courtney Nowell Date: 30SEP26

For Research Use Only