QUANSYS BIOSCIENCES

Quansys Biosciences, Inc. | 365 N 600 W Logan, UT 84321 | (435) 752-0531

CERTIFICATE OF ANALYSIS


Q-Plex™ Human Pneumococcal IgG (14-Plex) Aliquoted Bulk 30 mL Diluent

REF	Part Number: 482249HU-30-25	Software Product Code:
LOT	Lot Number: HPNK260113 24393 001	HPNM260113 23651 001
	Expiration Date: 2027-MAY-08	

Reconstitute Calibrator using 500 µL of Sample Diluent

Component	Part Number	Lot Number	Recon. Volume	Exp. Date
Plate	482252HU	HPNP260511	N/A	2027-MAY-11
Sample Diluent	481157HU-30	HPNS260928	N/A	2027-SEP-28
Calibrator	481155HU	HPNM260113	500 µL	2027-OCT-05
Detection	482156HU	HPND260508	N/A	2027-MAY-08
Substrate A	101159GR	QNSA260612	N/A	2028-APR-02
Substrate B+	101120GR	QNSB+260612	N/A	2028-APR-02
Wash Buffer	101158GR	QNSW260923	N/A	2028-SEP-23

Spot #	Analyte	Cal Conc	Units	LLD	LLOQ	ULOQ
1	1	1.41	µg/mL	0.00030	0.00060	0.38
2	3	0.36	µg/mL	0.00040	0.00079	0.35
3	4	0.84	µg/mL	0.00014	0.00027	0.28
4	5	1.89	µg/mL	0.00033	0.00066	0.83
5	6B	0.71	µg/mL	0.00012	0.00024	0.59
6	7F	2.09	µg/mL	0.00049	0.00098	0.55
7	8	0.64	µg/mL	0.00012	0.00024	0.33
8	9N	0.82	µg/mL	0.00014	0.00029	0.36
9	9V	1.06	µg/mL	0.00019	0.00039	0.28
10	12F	0.37	µg/mL	0.000063	0.00013	0.21
11	14	2.54	µg/mL	0.00054	0.0011	0.45
12	18C	1.07	µg/mL	0.00054	0.0011	0.24
13	19F	5.58	µg/mL	0.0021	0.0042	3.55
14	23F	0.42	µg/mL	0.00039	0.00077	0.28
15	CWPS	N/A	N/A	N/A	N/A	N/A
16	Control Spot	N/A	N/A	N/A	N/A	N/A
17	REF	N/A	N/A	N/A	N/A	N/A



Q-Plex™ ARRAYS

QC Release: Krzysztof Kowalczyk Date: 05OCT26

QA Release: Courtney Novak Date: 06OCT26

101896GR - 9.0

For Research Use Only